

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD061418\
 Data File : VD058151.D
 Acq On : 14 Jun 2018 22:31
 Operator : VA/AP
 Sample : J3241-11
 Misc : 5.47g/5ml/MSVOA D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-05-061318-C

Quant Time: Jun 15 07:48:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 08:57:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	629543	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.69	114	841874	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.69	117	534472	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.91	152	270419	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.02	65	260602	41.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.82%	
35) Dibromofluoromethane	5.55	113	277459	42.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.20%	
50) Toluene-d8	8.71	98	433115	27.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	55.94%	
62) 4-Bromofluorobenzene	11.93	95	232873	34.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.94%	
Target Compounds						
16) Acetone	2.62	43	24116	26.595	ug/l	97
20) Methylene Chloride	3.07	84	22829	3.033	ug/l	98
69) o-Xylene	11.41	106	6894	1.090	ug/l	95
80) 1,3,5-Trimethylbenzene	12.32	105	27546	1.972	ug/l	96
86) p-Isopropyltoluene	12.88	119	26831	1.749	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD061418\
Data File : VD058151.D
Acq On : 14 Jun 2018 22:31
Operator : VA/AP
Sample : J3241-11
Misc : 5.47g/5ml/MSVOA D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HR-05-061318-C

Quant Time: Jun 15 07:48:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 08:57:33 2018
Response via : Initial Calibration

