

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD061418\
 Data File : VD058136.D
 Acq On : 14 Jun 2018 15:27
 Operator : VA/AP
 Sample : J3420-05
 Misc : 4.66g/5ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-42(0-5)

Quant Time: Jun 15 07:02:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 08:57:33 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	1245627	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.70	114	1532296	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.70	117	1220229	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.91	152	587460	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.02	65	424444	34.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.16%	
35) Dibromofluoromethane	5.55	113	517630	43.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.32%	
50) Toluene-d8	8.71	98	1197691	42.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.98%	
62) 4-Bromofluorobenzene	11.94	95	467922	38.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.12%	
Target Compounds						
20) Methylene Chloride	3.08	84	65295	8.863	ug/l	Qvalue 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD061418\
Data File : VD058136.D
Acq On : 14 Jun 2018 15:27
Operator : VA/AP
Sample : J3420-05
Misc : 4.66g/5ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-42(0-5)

Quant Time: Jun 15 07:02:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 08:57:33 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 551

Delta R.T. -0.00 min

Lab File: VD058136.D

Acq: 14 Jun 2018 15:27

Instrument :

MSVOA_D

ClientSampleId :


SB-42(0-5)

Tot Ion:168 Resp: 1245627

Ion Ratio Lower Upper




168 100

99 53.4 46.8 70.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC



#20

Methylene Chloride

Concen: 8.863 ug/l

RT: 3.08 min Scan# 290

Delta R.T. 0.00 min

Lab File: VD058136.D

Acq: 14 Jun 2018 15:27

Tgt Ion: 84 Resp: 65295


Ion Ratio Lower Upper

84 100

49 173.7 141.4 212.0

51 50.7 38.3 57.5

86 62.9 52.9 79.3






Abundance Ion 83.95 (83.65 to 84.65): VDC

Ion 48.90 (48.60 to 49.60): VDC

Ion 50.90 (50.60 to 51.60): VDC

Ion 85.90 (85.60 to 86.60): VDC









#33
 1,2-Dichloroethane-d4
 Concen: N.D. ug/l
 RT: 6.02 min Scan# 589
 Delta R.T. -0.00 min
 Lab File: VD058136.D
 Acq: 14 Jun 2018 15:27




Instrument :
 MSVOA_D
 ClientSampleId :
 SB-42(0-5)


Tot Ion: 65 Resp: 424444
 Ion Ratio Lower Upper
 65 100
 67 48.6 0.0 94.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.70 min Scan# 658
 Delta R.T. -0.00 min
 Lab File: VD058136.D
 Acq: 14 Jun 2018 15:27

Tgt Ion: 114 Resp: 1532296
 Ion Ratio Lower Upper
 114 100
 63 21.5 0.0 42.8
 88 18.4 0.0 37.8





#35

Dibromofluoromethane

Concen: 50.000 ug/l

RT: 5.55 min Scan# 541

Delta R.T. -0.01 min

Lab File: VD058136.D

Acq: 14 Jun 2018 15:27

Instrument :

MSVOA_D

ClientSampleId :

SB-42(0-5)


Tot Ion:113 Resp: 517630

Ion Ratio Lower Upper

113 100

111 102.6 81.5 122.3




192 20.6 16.4 24.6



Abundance Ion 112.90 (112.60 to 113.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 191.80 (191.50 to 192.50): V



#50

Toluene-d8

Concen: N.D. ug/l

RT: 8.71 min Scan# 862

Delta R.T. -0.00 min

Lab File: VD058136.D


Acq: 14 Jun 2018 15:27

Tgt Ion: 98 Resp: 1197691

Ion Ratio Lower Upper



98 100


100 68.6 53.4 80.0



Abundance Ion 98.05 (97.75 to 98.75): V

Ion 100.05 (99.75 to 100.75): V





#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 11.94 min Scan# 1190

Delta R.T. 0.01 min

Lab File: VD058136.D

Acq: 14 Jun 2018 15:27

Instrument :

MSVOA_D

ClientSampleId :

SB-42(0-5)


Tot Ion: 95 Resp: 467922

Ion Ratio Lower Upper

95 100

174 100.5 0.0 169.4




176 100.4 0.0 162.6



Abundance Ion 95.00 (94.70 to 95.70): V

Ion 173.90 (173.60 to 174.60): V

Ion 175.90 (175.60 to 176.60): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.70 min Scan# 1064

Delta R.T. 0.01 min

Lab File: VD058136.D

Acq: 14 Jun 2018 15:27


Tgt Ion: 117 Resp: 1220229

Ion Ratio Lower Upper

117 100

82 60.4 46.8 70.2



119 31.9 25.8 38.6




Abundance Ion 116.95 (116.65 to 117.65): V

Ion 82.05 (81.75 to 82.75): V

Ion 118.95 (118.65 to 119.65): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.91 min Scan# 1289

Delta R.T. -0.00 min

Lab File: VD058136.D

Acq: 14 Jun 2018 15:27

Instrument :

MSVOA_D

ClientSampleId :

SB-42(0-5)


Tot Ion:152 Resp: 587460

Ion Ratio Lower Upper

152 100

115 54.6 25.6 76.6

150 152.7 0.0 306.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 115.00 (114.70 to 115.70): V

Ion 149.90 (149.60 to 150.60): V

