

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD061518\
 Data File : VD058163.D
 Acq On : 15 Jun 2018 20:25
 Operator : VA/AP
 Sample : J3493-01
 Misc : 5.02g/5ml/MSVOA D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1

Quant Time: Jun 16 00:20:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 08:57:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	47584	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.69	114	72406	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.70	117	63481	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.92	152	28041	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.02	65	71903	151.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	302.30%	
35) Dibromofluoromethane	5.55	113	49670	88.66	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	177.32%	
50) Toluene-d8	8.72	98	85759	64.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.78%	
62) 4-Bromofluorobenzene	11.95	95	37907	65.24	ug/l	0.02
Spiked Amount 50.000			Recovery	=	130.48%	
Target Compounds						
16) Acetone	2.60	43	16203	236.407	ug/l	92
27) cis-1,2-Dichloroethene	4.80	96	149048	295.647	ug/l	86
44) Trichloroethene	6.99	130	125061	229.753	ug/l	100
64) Tetrachloroethene	9.53	164	702964	1238.658	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD061518\
Data File : VD058163.D
Acq On : 15 Jun 2018 20:25
Operator : VA/AP
Sample : J3493-01
Misc : 5.02g/5ml/MSVOA D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1

Quant Time: Jun 16 00:20:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 08:57:33 2018
Response via : Initial Calibration

