

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061622\
 Data File : VD073614.D
 Acq On : 16 Jun 2022 13:37
 Operator : VA/SY
 Sample : N3324-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BT-SPOILS-07

Quant Time: Jun 17 01:05:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	281107	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	483998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	445343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	189771	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	159322	52.814	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.620%
35) Dibromofluoromethane	7.909	113	160365	51.526	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.060%
50) Toluene-d8	10.332	98	581981	49.458	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.920%
62) 4-Bromofluorobenzene	12.620	95	181459	42.478	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061622\
Data File : VD073614.D
Acq On : 16 Jun 2022 13:37
Operator : VA/SY
Sample : N3324-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BT-SPOILS-07

Quant Time: Jun 17 01:05:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

