

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061622\
 Data File : VD073618.D
 Acq On : 16 Jun 2022 15:33
 Operator : VA/SY
 Sample : N3339-02
 Misc : 7.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CB-226-(26-28)

Quant Time: Jun 17 01:05:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	177015	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	376767	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	225942	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	18614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	153008	80.547	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	161.100%	
35) Dibromofluoromethane	7.909	113	147327	60.810	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	121.620%	
50) Toluene-d8	10.332	98	403538	44.054	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	88.100%	
62) 4-Bromofluorobenzene	12.620	95	47442	14.267	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	28.540%	
Target Compounds						Qvalue
17) Carbon Disulfide	4.409	76	24008	4.385	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061622\
Data File : VD073618.D
Acq On : 16 Jun 2022 15:33
Operator : VA/SY
Sample : N3339-02
Misc : 7.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CB-226-(26-28)

Quant Time: Jun 17 01:05:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

