

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061622\
 Data File : VD073628.D
 Acq On : 16 Jun 2022 20:23
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 17 01:07:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	74	0.00
2 T	Dichlorodifluoromethane	0.477	0.426	10.7	69	0.00
3 P	Chloromethane	0.489	0.417	14.7	66	0.00
4 C	Vinyl Chloride	0.459	0.390	15.0#	63	0.00
5 T	Bromomethane	0.329	0.283	14.0	67	0.00
6 T	Chloroethane	0.292	0.253	13.4	64	0.00
7 T	Trichlorofluoromethane	0.898	0.923	-2.8	76	0.00
8 T	Diethyl Ether	0.257	0.261	-1.6	77	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.514	0.538	-4.7	75	0.00
10 T	Methyl Iodide	0.536	0.568	-6.0	73	0.00
11 T	Tert butyl alcohol	0.150	0.055	63.3#	63	0.00
12 CM	1,1-Dichloroethene	0.488	0.487	0.2#	73	0.00
13 T	Acrolein	0.024	0.015	37.5#	53	0.00
14 T	Allyl chloride	0.726	0.730	-0.6	76	0.00
15 T	Acrylonitrile	0.115	0.121	-5.2	79	0.00
16 T	Acetone	0.094	0.090	4.3	77	0.00
17 T	Carbon Disulfide	1.546	1.453	6.0	67	0.00
18 T	Methyl Acetate	0.285	0.259	9.1	81	0.00
19 T	Methyl tert-butyl Ether	1.164	1.283	-10.2	81	0.00
20 T	Methylene Chloride	0.668	0.686	-2.7	84	0.00
21 T	trans-1,2-Dichloroethene	0.590	0.624	-5.8	76	0.00
22 T	Diisopropyl ether	1.567	1.702	-8.6	78	0.00
23 T	Vinyl Acetate	0.850	0.944	-11.1	78	0.00
24 P	1,1-Dichloroethane	1.005	1.062	-5.7	78	0.00
25 T	2-Butanone	0.139	0.139	0.0	78	0.00
26 T	2,2-Dichloropropane	0.901	0.907	-0.7	75	0.00
27 T	cis-1,2-Dichloroethene	0.638	0.703	-10.2	80	0.00
28 T	Bromochloromethane	0.394	0.400	-1.5	76	0.00
29 T	Tetrahydrofuran	0.089	0.095	-6.7	78	0.00
30 C	Chloroform	1.073	1.218	-13.5#	84	0.00
31 T	Cyclohexane	0.904	0.857	5.2	69	0.00
32 T	1,1,1-Trichloroethane	0.946	1.049	-10.9	80	0.00
33 S	1,2-Dichloroethane-d4	0.537	0.634	-18.1	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
35 S	Dibromofluoromethane	0.322	0.382	-18.6	89	0.00
36 T	1,1-Dichloropropene	0.483	0.504	-4.3	75	0.00
37 T	Ethyl Acetate	0.195	0.192	1.5	73	0.00
38 T	Carbon Tetrachloride	0.505	0.564	-11.7	81	0.00
39 T	Methylcyclohexane	0.563	0.564	-0.2	69	0.00
40 TM	Benzene	1.405	1.488	-5.9	77	0.00
41 T	Methacrylonitrile	0.106	0.102	3.8	67	0.00
42 TM	1,2-Dichloroethane	0.388	0.445	-14.7	85	0.00
43 T	Isopropyl Acetate	0.375	0.385	-2.7	79	0.00
44 TM	Trichloroethene	0.391	0.406	-3.8	76	0.00
45 C	1,2-Dichloropropane	0.348	0.371	-6.6#	80	0.00
46 T	Dibromomethane	0.196	0.212	-8.2	83	0.00
47 T	Bromodichloromethane	0.491	0.556	-13.2	84	0.00
48 T	Methyl methacrylate	0.186	0.202	-8.6	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061622\
 Data File : VD073628.D
 Acq On : 16 Jun 2022 20:23
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 17 01:07:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	84	0.00
50 S	Toluene-d8	1.216	1.412	-16.1	84	0.00
51 T	4-Methyl-2-Pentanone	0.190	0.206	-8.4	81	0.00
52 CM	Toluene	0.901	1.003	-11.3#	79	0.00
53 T	t-1,3-Dichloropropene	0.461	0.512	-11.1	82	0.00
54 T	cis-1,3-Dichloropropene	0.540	0.596	-10.4	82	0.00
55 T	1,1,2-Trichloroethane	0.280	0.302	-7.9	82	0.00
56 T	Ethyl methacrylate	0.320	0.357	-11.6	82	0.00
57 T	1,3-Dichloropropane	0.455	0.503	-10.5	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.132	0.138	-4.5	77	0.00
59 T	2-Hexanone	0.130	0.140	-7.7	81	0.00
60 T	Dibromochloromethane	0.345	0.394	-14.2	84	0.00
61 T	1,2-Dibromoethane	0.266	0.294	-10.5	83	0.00
62 S	4-Bromofluorobenzene	0.441	0.527	-19.5	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	78	0.00
64 T	Tetrachloroethene	0.340	0.352	-3.5	76	0.00
65 PM	Chlorobenzene	1.029	1.100	-6.9	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.390	0.426	-9.2	81	0.00
67 C	Ethyl Benzene	1.798	1.963	-9.2#	78	0.00
68 T	m/p-Xylenes	0.706	0.777	-10.1	79	0.00
69 T	o-Xylene	0.649	0.727	-12.0	80	0.00
70 T	Styrene	1.120	1.276	-13.9	81	0.00
71 P	Bromoform	0.213	0.240	-12.7	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.466	3.750	-8.2	79	0.00
74 T	N-amyl acetate	0.771	0.787	-2.1	82	0.00
75 P	1,1,2,2-Tetrachloroethane	0.658	0.688	-4.6	85	0.00
76 T	1,2,3-Trichloropropane	0.486	0.487	-0.2	80	0.00
77 T	Bromobenzene	0.846	0.907	-7.2	84	0.00
78 T	n-propylbenzene	4.314	4.639	-7.5	79	0.00
79 T	2-Chlorotoluene	2.508	2.682	-6.9	81	0.00
80 T	1,3,5-Trimethylbenzene	2.989	3.283	-9.8	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.207	0.217	-4.8	83	0.00
82 T	4-Chlorotoluene	2.644	2.829	-7.0	81	0.00
83 T	tert-Butylbenzene	2.526	2.815	-11.4	80	0.00
84 T	1,2,4-Trimethylbenzene	2.983	3.295	-10.5	82	0.00
85 T	sec-Butylbenzene	3.784	4.125	-9.0	79	0.00
86 T	p-Isopropyltoluene	3.155	3.496	-10.8	81	0.00
87 T	1,3-Dichlorobenzene	1.728	1.817	-5.2	82	0.00
88 T	1,4-Dichlorobenzene	1.710	1.822	-6.5	83	0.00
89 T	n-Butylbenzene	2.976	3.233	-8.6	79	0.00
90 T	Hexachloroethane	0.653	0.692	-6.0	83	0.00
91 T	1,2-Dichlorobenzene	1.491	1.643	-10.2	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.105	0.107	-1.9	86	0.00
93 T	1,2,4-Trichlorobenzene	0.893	0.956	-7.1	82	0.00
94 T	Hexachlorobutadiene	0.490	0.522	-6.5	80	0.00
95 T	Naphthalene	1.596	1.778	-11.4	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061622\
Data File : VD073628.D
Acq On : 16 Jun 2022 20:23
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jun 17 01:07:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.779	0.828	-6.3	81	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6