

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061622\
 Data File : VD073627.D
 Acq On : 16 Jun 2022 19:54
 Operator : VA/SY
 Sample : N3339-02RE
 Misc : 7.52G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CB-226-(26-28)RE

Quant Time: Jun 17 01:07:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	255979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	458540	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	390760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	94786	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	155697	56.679	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.360%
35) Dibromofluoromethane	7.908	113	154131	52.273	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.540%
50) Toluene-d8	10.332	98	549157	49.259	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.520%
62) 4-Bromofluorobenzene	12.620	95	140401	34.692	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	69.380%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061622\
Data File : VD073627.D
Acq On : 16 Jun 2022 19:54
Operator : VA/SY
Sample : N3339-02RE
Misc : 7.52G/5.00ml/MSVOA_D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CB-226-(26-28)RE

Quant Time: Jun 17 01:07:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.973 min Scan# 1029
Delta R.T. -0.000 min
Lab File: VD073627.D
Acq: 16 Jun 2022 19:54




Instrument :
MSVOA_D
ClientSampleId :
CB-226-(26-28)RE


Tgt Ion:168 Resp: 255979
Ion Ratio Lower Upper
168 100
99 62.4 49.8 74.6



#33
1,2-Dichloroethane-d4
Concen: 56.679 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. 0.006 min
Lab File: VD073627.D
Acq: 16 Jun 2022 19:54

Tgt Ion: 65 Resp: 155697
Ion Ratio Lower Upper
65 100
67 52.7 0.0 101.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.855 min Scan# 1179

Delta R.T. 0.000 min

Lab File: VD073627.D

Acq: 16 Jun 2022 19:54

Instrument :

MSVOA_D

ClientSampleId :

CB-226-(26-28)RE





Tgt Ion:114 Resp: 458540

Ion Ratio Lower Upper

114 100

63 20.1 0.0 46.6

88 15.6 0.0 33.2



#35

Dibromofluoromethane

Concen: 52.273 ug/l

RT: 7.908 min Scan# 1018

Delta R.T. 0.000 min

Lab File: VD073627.D

Acq: 16 Jun 2022 19:54




Tgt Ion:113 Resp: 154131


Ion Ratio Lower Upper

113 100

111 104.6 83.8 125.6

192 19.0 14.6 21.8









#50
Toluene-d8
Concen: 49.259 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD073627.D
Acq: 16 Jun 2022 19:54




Instrument :
MSVOA_D
ClientSampleId :
CB-226-(26-28)RE


Tgt Ion: 98 Resp: 549157
Ion Ratio Lower Upper
98 100
100 65.3 52.4 78.6



#62
4-Bromofluorobenzene
Concen: 34.692 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. -0.000 min
Lab File: VD073627.D
Acq: 16 Jun 2022 19:54

Tgt Ion: 95 Resp: 140401
Ion Ratio Lower Upper
95 100
174 80.2 0.0 152.8
176 77.7 0.0 145.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.637 min Scan# 1

Delta R.T. 0.005 min

Lab File: VD073627.D

Acq: 16 Jun 2022 19:54

Instrument :

MSVOA_D

ClientSampleId :

CB-226-(26-28)RE





Tgt Ion:117 Resp: 390760

Ion Ratio Lower Upper

117 100

82 54.4 45.3 67.9

119 32.9 26.6 40.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.561 min Scan# 1979

Delta R.T. -0.000 min

Lab File: VD073627.D

Acq: 16 Jun 2022 19:54

Tgt Ion:152 Resp: 94786

Ion Ratio Lower Upper

152 100

115 60.7 31.3 93.8

150 160.1 0.0 348.6

