

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061824\
 Data File : VD079200.D
 Acq On : 18 Jun 2024 15:21
 Operator : RP/MD
 Sample : P2897-12
 Misc : 3.60G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COT59

Quant Time: Jun 19 07:18:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jun 19 07:16:52 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		8.775	114	139204	25.000	ug/L	0.00
28) Chlorobenzene-d5		11.581	117	126361	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		13.516	152	33627	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		2.270	65	136409	31.844	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	127.360%	
7) Chloroethane-d5		2.799	69	120518	29.096	ug/L	0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	116.400%	
11) 1,1-Dichloroethene-d2		3.905	65	26335	34.630	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	138.520%#	
21) 2-Butanone-d5		6.981	46	27099	82.707	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	165.420%#	
24) Chloroform-d		7.563	84	115956	29.792	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	119.160%	
26) 1,2-Dichloroethane-d4		8.222	65	64253	35.785	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	143.160%#	
32) Benzene-d6		8.199	84	236789	33.803	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	135.200%#	
36) 1,2-Dichloropropane-d6		9.204	67	71219	35.295	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	141.160%#	
41) Toluene-d8		10.263	98	194612	29.672	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	118.680%	
43) trans-1,3-Dichloroprop...		10.516	79	30348	35.782	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	143.120%#	
47) 2-Hexanone-d5		10.875	63	23570	90.355	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	180.720%#	
56) 1,1,2,2-Tetrachloroeth...		12.645	84	64423	36.045	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	144.200%#	
66) 1,2-Dichlorobenzene-d4		13.810	152	35788	30.838	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	123.360%#	
Target Compounds							
16) Methylene chloride		4.787	84	30963	12.483	ug/L	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061824\
Data File : VD079200.D
Acq On : 18 Jun 2024 15:21
Operator : RP/MD
Sample : P2897-12
Misc : 3.60G/10ml/MSVOA_D/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0T59

Quant Time: Jun 19 07:18:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jun 19 07:16:52 2024
Response via : Initial Calibration

