

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061923\
 Data File : VD076468.D
 Acq On : 19 Jun 2023 14:11
 Operator : KP/SY
 Sample : 03262-01
 Misc : 5.13G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WS-SOIL1-0615

Quant Time: Jun 20 05:24:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	229428	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	397115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	340334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	137347	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	107801	46.897	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.800%
35) Dibromofluoromethane	7.804	113	117279	49.713	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.420%
50) Toluene-d8	10.269	98	442184	49.423	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.575	95	124254	42.792	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061923\
Data File : VD076468.D
Acq On : 19 Jun 2023 14:11
Operator : KP/SY
Sample : 03262-01
Misc : 5.13G/5.0ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WS-SOIL1-0615

Quant Time: Jun 20 05:24:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

