

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061923\
 Data File : VD076472.D
 Acq On : 19 Jun 2023 16:12
 Operator : KP/SY
 Sample : 03289-01
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-01-0-2.0

Quant Time: Jun 20 05:25:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	197411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	337585	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	306981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	136351	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	99823	50.469	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.940%	
35) Dibromofluoromethane	7.804	113	101099	50.411	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.820%	
50) Toluene-d8	10.269	98	379946	49.955	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 99.920%	
62) 4-Bromofluorobenzene	12.569	95	115321	46.719	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 93.440%	
Target Compounds						
16) Acetone	4.028	43	10677	22.685	ug/l	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061923\
Data File : VD076472.D
Acq On : 19 Jun 2023 16:12
Operator : KP/SY
Sample : 03289-01
Misc : 5.01G/5.0ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-01-0-2.0

Quant Time: Jun 20 05:25:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

