

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061923\
 Data File : VD076473.D
 Acq On : 19 Jun 2023 16:43
 Operator : KP/SY
 Sample : 03289-02
 Misc : 5.09G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-01-4.0-6.0

Quant Time: Jun 20 05:26:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	3745	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	6899	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	7221	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	3823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	3570	95.145	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 190.280%#	
35) Dibromofluoromethane	7.805	113	2537	61.901	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 123.800%	
50) Toluene-d8	10.269	98	8774	56.449	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 112.900%	
62) 4-Bromofluorobenzene	12.569	95	2892	57.330	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 114.660%	
Target Compounds						Qvalue
3) Chloromethane	2.140	50	703	11.878	ug/l	98
5) Bromomethane	2.705	94	465	10.833	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061923\
Data File : VD076473.D
Acq On : 19 Jun 2023 16:43
Operator : KP/SY
Sample : 03289-02
Misc : 5.09G/5.0ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-01-4.0-6.0

Quant Time: Jun 20 05:26:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

