

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061923\
 Data File : VD076476.D
 Acq On : 19 Jun 2023 18:14
 Operator : KP/SY
 Sample : 03289-07
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-03-0-2.0

Quant Time: Jun 20 05:27:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	3911	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	7307	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	6459	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	3049	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.240	65	3084	78.704	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	157.400%
35) Dibromofluoromethane	7.799	113	2800	64.503	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	129.000%
50) Toluene-d8	10.269	98	7417	45.054	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.100%
62) 4-Bromofluorobenzene	12.575	95	3050	57.086	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	114.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061923\
Data File : VD076476.D
Acq On : 19 Jun 2023 18:14
Operator : KP/SY
Sample : 03289-07
Misc : 5.01G/5.0ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-03-0-2.0

Quant Time: Jun 20 05:27:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

