

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061923\
 Data File : VD076479.D
 Acq On : 19 Jun 2023 19:46
 Operator : KP/SY
 Sample : 03289-13
 Misc : 5.06G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-05-12.0-14.0

Quant Time: Jun 20 05:28:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	1503	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.769	114	3374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	3446	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	1095	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	1492	99.078	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	198.160%#
35) Dibromofluoromethane	7.799	113	467	23.299	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	46.600%#
50) Toluene-d8	10.275	98	3946	51.910	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.820%
62) 4-Bromofluorobenzene	12.569	95	1410	57.154	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	114.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061923\
Data File : VD076479.D
Acq On : 19 Jun 2023 19:46
Operator : KP/SY
Sample : 03289-13
Misc : 5.06G/5.0ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-05-12.0-14.0

Quant Time: Jun 20 05:28:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

