



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061923\
Data File : VD076463.D
Acq On : 19 Jun 2023 08:46
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Title : SW846 8260
Last Update : Mon Jun 19 04:51:45 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	21.7	12421	PASS
75	95	30	60	52.7	30187	PASS
95	95	100	100	100.0	57259	PASS
96	95	5	9	6.8	3916	PASS
173	174	0.00	2	1.1	551	PASS
174	95	50	100	87.3	50008	PASS
175	174	5	9	8.5	4236	PASS
176	174	95	101	99.7	49861	PASS
177	176	5	9	6.9	3446	PASS