


Dichlorodifluoromethane

Response Ratio




$$\text{Response} = 4.000\text{e-}001 * \text{Amt} + 2.986\text{e-}002$$

Coef of Det (r^2) = 0.999206 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D062022S.M

Calibration Table Last Updated: Tue Jun 21 05:24:45 2022


Bromomethane



Response = 2.720e-001 * Amt + 1.566e-002
 Coef of Det (r^2) = 0.999605 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D062022S.M
 Calibration Table Last Updated: Tue Jun 21 05:24:45 2022

Naphthalene

Response Ratio




$$\text{Response} = 1.864\text{e}+000 * \text{Amt} - 1.002\text{e}-001$$

Coef of Det (r^2) = 0.998760 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D062022S.M


Calibration Table Last Updated: Tue Jun 21 05:24:45 2022

Tert butyl alcohol




Response = $3.018 \times 10^{-2} \times \text{Amt} + 5.632 \times 10^{-2}$
Coef of Det (r^2) = 0.992479 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D062022S.M
Calibration Table Last Updated: Tue Jun 21 05:24:45 2022

Acrolein



$\text{Response} = 2.309\text{e-}002 * \text{Amt} - 9.725\text{e-}003$
 Coef of Det (r^2) = 0.996776 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D062022S.M
 Calibration Table Last Updated: Tue Jun 21 05:24:45 2022


Methyl Acetate



Response = 2.365e-001 * Amt + 1.419e-002
Coef of Det (r^2) = 0.999329 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D062022S.M
Calibration Table Last Updated: Tue Jun 21 05:24:45 2022

Methylene Chloride

Response Ratio



$$\text{Response} = 5.360\text{e-}001 * \text{Amt} + 6.442\text{e-}002$$

Coef of Det (r^2) = 0.998055 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D062022S.M

Calibration Table Last Updated: Tue Jun 21 05:24:45 2022