

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062024\
 Data File : VD079220.D
 Acq On : 20 Jun 2024 16:45
 Operator : RP/MD
 Sample : VIBLK807
 Misc : 5.32G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COT67

Quant Time: Jun 21 04:53:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 21 04:13:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/L	-8.78
28) Chlorobenzene-d5	0.000	117	0	0.000	ug/L	-11.58
58) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/L	-13.52
System Monitoring Compounds						
4) Vinyl Chloride-d3	0.000	65	0	0.000	ug/L	
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.000%#
7) Chloroethane-d5	0.000	69	0	0.000	ug/L	
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.000%#
11) 1,1-Dichloroethene-d2	0.000	65	0	0.000	ug/L	
Spiked Amount	25.000	Range	45 - 110	Recovery	=	0.000%#
21) 2-Butanone-d5	0.000	46	0	0.000	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
24) Chloroform-d	0.000	84	0	0.000	ug/L	
Spiked Amount	25.000	Range	40 - 150	Recovery	=	0.000%#
26) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/L	
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.000%#
32) Benzene-d6	0.000	84	0	0.000	ug/L	
Spiked Amount	25.000	Range	20 - 135	Recovery	=	0.000%#
36) 1,2-Dichloropropane-d6	0.000	67	0	0.000	ug/L	
Spiked Amount	25.000	Range	70 - 120	Recovery	=	0.000%#
41) Toluene-d8	0.000	98	0	0.000	ug/L	
Spiked Amount	25.000	Range	30 - 130	Recovery	=	0.000%#
43) trans-1,3-Dichloroprop...	0.000	79	0	0.000	ug/L	
Spiked Amount	25.000	Range	30 - 135	Recovery	=	0.000%#
47) 2-Hexanone-d5	0.000	63	0	0.000	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
56) 1,1,2,2-Tetrachloroeth...	0.000	84	0	0.000	ug/L	
Spiked Amount	25.000	Range	45 - 120	Recovery	=	0.000%#
66) 1,2-Dichlorobenzene-d4	0.000	152	0	0.000	ug/L	
Spiked Amount	25.000	Range	75 - 120	Recovery	=	0.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062024\
Data File : VD079220.D
Acq On : 20 Jun 2024 16:45
Operator : RP/MD
Sample : VIBLK807
Misc : 5.32G/10ml/MSVOA_D/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0T67

Quant Time: Jun 21 04:53:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 21 04:13:10 2024
Response via : Initial Calibration

