

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062024\
 Data File : VD079217.D
 Acq On : 20 Jun 2024 15:22
 Operator : RP/MD
 Sample : P2897-16
 Misc : 3.46G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C0T63

Quant Time: Jun 21 04:14:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 21 04:13:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	306991	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	300927	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	131410	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.269	65	123294	13.051	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	52.200%
7) Chloroethane-d5	2.793	69	118241	12.944	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	51.760%
11) 1,1-Dichloroethene-d2	3.911	65	22011	13.125	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	52.480%
21) 2-Butanone-d5	6.981	46	28967	40.089	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.180%
24) Chloroform-d	7.563	84	130237	15.173	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	60.680%
26) 1,2-Dichloroethane-d4	8.228	65	66308	16.746	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	67.000%#
32) Benzene-d6	8.199	84	260124	15.593	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	62.360%
36) 1,2-Dichloropropane-d6	9.204	67	78981	16.436	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	65.760%#
41) Toluene-d8	10.263	98	234121	14.989	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	59.960%
43) trans-1,3-Dichloroprop...	10.522	79	32785	16.232	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	64.920%
47) 2-Hexanone-d5	10.875	63	23379	37.633	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.260%
56) 1,1,2,2-Tetrachloroeth...	12.645	84	74289	17.453	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	69.800%
66) 1,2-Dichlorobenzene-d4	13.810	152	83871	18.493	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	73.960%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062024\
Data File : VD079217.D
Acq On : 20 Jun 2024 15:22
Operator : RP/MD
Sample : P2897-16
Misc : 3.46G/10ml/MSVOA_D/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0T63

Quant Time: Jun 21 04:14:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 21 04:13:10 2024
Response via : Initial Calibration

