

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062024\
 Data File : VD079218.D
 Acq On : 20 Jun 2024 15:50
 Operator : RP/MD
 Sample : P2897-17
 Misc : 4.15G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C0T64

Quant Time: Jun 21 04:15:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 21 04:13:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	219600	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	201359	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	62730	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	104892	15.522	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.080%
7) Chloroethane-d5	2.793	69	105264	16.110	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.440%
11) 1,1-Dichloroethene-d2	3.911	65	18182	15.156	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.640%
21) 2-Butanone-d5	6.981	46	26637	51.534	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.060%
24) Chloroform-d	7.563	84	102353	16.670	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	66.680%
26) 1,2-Dichloroethane-d4	8.228	65	54488	19.237	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.960%
32) Benzene-d6	8.199	84	198846	17.814	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	71.240%
36) 1,2-Dichloropropane-d6	9.204	67	61548	19.141	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	76.560%
41) Toluene-d8	10.269	98	167024	15.981	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	63.920%
43) trans-1,3-Dichloroprop...	10.522	79	24789	18.342	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.360%
47) 2-Hexanone-d5	10.875	63	20538	49.408	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.820%
56) 1,1,2,2-Tetrachloroeth...	12.645	84	61482	21.587	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	86.360%
66) 1,2-Dichlorobenzene-d4	13.810	152	40164	18.552	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	74.200%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062024\
Data File : VD079218.D
Acq On : 20 Jun 2024 15:50
Operator : RP/MD
Sample : P2897-17
Misc : 4.15G/10ml/MSVOA_D/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0T64

Quant Time: Jun 21 04:15:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 21 04:13:10 2024
Response via : Initial Calibration

