

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062222\
 Data File : VD073659.D
 Acq On : 22 Jun 2022 17:19
 Operator : VA/SY
 Sample : N3421-04
 Misc : 3.35G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3-0-2-20220620

Quant Time: Jun 23 09:29:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 21 05:24:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	237373	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	435194	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	401562	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	177821	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	139343	51.018	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.040%
35) Dibromofluoromethane	7.908	113	145246	46.006	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.020%
50) Toluene-d8	10.332	98	518286	45.867	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.740%
62) 4-Bromofluorobenzene	12.620	95	166165	40.098	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062222\
Data File : VD073659.D
Acq On : 22 Jun 2022 17:19
Operator : VA/SY
Sample : N3421-04
Misc : 3.35G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-3-0-2-20220620

Quant Time: Jun 23 09:29:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062022S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 21 05:24:45 2022
Response via : Initial Calibration

