

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062222\
 Data File : VD073661.D
 Acq On : 22 Jun 2022 18:16
 Operator : VA/SY
 Sample : N3421-06
 Misc : 4.17G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3-23-25-20220620

Quant Time: Jun 23 09:29:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 21 05:24:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	228334	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	416292	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	394552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	177257	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	139764	53.198	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.400%
35) Dibromofluoromethane	7.908	113	144003	47.684	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.360%
50) Toluene-d8	10.332	98	508666	47.059	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.120%
62) 4-Bromofluorobenzene	12.620	95	160936	40.600	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062222\
Data File : VD073661.D
Acq On : 22 Jun 2022 18:16
Operator : VA/SY
Sample : N3421-06
Misc : 4.17G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-3-23-25-20220620

Quant Time: Jun 23 09:29:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062022S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 21 05:24:45 2022
Response via : Initial Calibration

