

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070324\
 Data File : VD079333.D
 Acq On : 03 Jul 2024 14:10
 Operator : RP/MD
 Sample : P3149-01RE
 Misc : 5.06G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-01-070224RE

Quant Time: Jul 04 06:23:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 29 06:28:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	21890	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	50036	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	51245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	16880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	25936	127.504	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 255.000%#	
35) Dibromofluoromethane	7.804	113	23169	70.544	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 141.080%	
50) Toluene-d8	10.269	98	43989	37.230	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 74.460%	
62) 4-Bromofluorobenzene	12.569	95	16427	43.356	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 86.720%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.793	84	8205	19.570	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070324\
Data File : VD079333.D
Acq On : 03 Jul 2024 14:10
Operator : RP/MD
Sample : P3149-01RE
Misc : 5.06G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-01-070224RE

Quant Time: Jul 04 06:23:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 29 06:28:31 2024
Response via : Initial Calibration

