

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070824\
 Data File : VD079348.D
 Acq On : 08 Jul 2024 18:05
 Operator : RP/MD
 Sample : VD0708SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0708SBL01

Quant Time: Jul 08 18:58:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070824S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 17:06:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	166021	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	332583	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	306030	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	125924	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	111411	56.945	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.900%
35) Dibromofluoromethane	7.804	113	133242	56.788	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.580%
50) Toluene-d8	10.263	98	489681	54.141	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.280%
62) 4-Bromofluorobenzene	12.569	95	140650	50.033	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	100.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070824\
Data File : VD079348.D
Acq On : 08 Jul 2024 18:05
Operator : RP/MD
Sample : VD0708SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0708SBL01

Quant Time: Jul 08 18:58:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070824S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 17:06:59 2024
Response via : Initial Calibration

