

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070824\
 Data File : VD079350.D
 Acq On : 08 Jul 2024 19:29
 Operator : RP/MD
 Sample : P3166-01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NB-08-070324

Quant Time: Jul 08 19:25:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070824S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 17:06:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	141133	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	290818	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	261960	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	91409	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	102856	61.843	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.680%	
35) Dibromofluoromethane	7.804	113	114030	55.580	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.160%	
50) Toluene-d8	10.269	98	399527	50.517	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.040%	
62) 4-Bromofluorobenzene	12.569	95	103782	42.220	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 84.440%	
Target Compounds						Qvalue
20) Methylene Chloride	4.799	84	33889	6.282	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070824\
Data File : VD079350.D
Acq On : 08 Jul 2024 19:29
Operator : RP/MD
Sample : P3166-01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NB-08-070324

Quant Time: Jul 08 19:25:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070824S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 17:06:59 2024
Response via : Initial Calibration

