

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070824\
 Data File : VD079356.D
 Acq On : 08 Jul 2024 22:15
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jul 09 05:01:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070824S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 17:06:59 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024
 Supervised By :Semsettin Yesilyurt 07/09/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	176473	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	316296	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	302348	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	150004	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	105755	50.853	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 101.700%			
35) Dibromofluoromethane	7.805	113	115257	51.652	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 103.300%			
50) Toluene-d8	10.269	98	454344	52.821	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 105.640%			
62) 4-Bromofluorobenzene	12.569	95	145651	54.480	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 108.960%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.928	85	77865	47.589	ug/l	98
3) Chloromethane	2.146	50	102143	46.624	ug/l	97
4) Vinyl Chloride	2.275	62	101698	45.167	ug/l	98
5) Bromomethane	2.681	94	85249	59.752	ug/l	98
6) Chloroethane	2.828	64	76980	63.808	ug/l	100
7) Trichlorofluoromethane	3.164	101	164773	50.083	ug/l	97
8) Diethyl Ether	3.593	74	52525	49.975	ug/l	89
9) 1,1,2-Trichlorotrifluo...	3.958	101	104471	50.625	ug/l	97
10) Methyl Iodide	4.158	142	123242	54.143	ug/l	98
11) Tert butyl alcohol	5.064	59	29054m	218.860	ug/l	
12) 1,1-Dichloroethene	3.934	96	101596	49.574	ug/l	99
13) Acrolein	3.799	56	28931	175.788	ug/l	96
14) Allyl chloride	4.558	41	127203	48.707	ug/l	87
15) Acrylonitrile	5.252	53	126517	261.335	ug/l	96
16) Acetone	4.022	43	92877	234.818	ug/l	100
17) Carbon Disulfide	4.264	76	322395	49.794	ug/l	96
18) Methyl Acetate	4.558	43	62712	50.673	ug/l	89
19) Methyl tert-butyl Ether	5.316	73	240074	52.711	ug/l	98
20) Methylene Chloride	4.793	84	179679	59.704	ug/l	97
21) trans-1,2-Dichloroethene	5.311	96	117897	51.787	ug/l	89
22) Diisopropyl ether	6.216	45	312958	53.514	ug/l	96
23) Vinyl Acetate	6.152	43	1188930	265.654	ug/l	95
24) 1,1-Dichloroethane	6.111	63	202734	51.414	ug/l	95
25) 2-Butanone	7.081	43	140358	249.828	ug/l	96
26) 2,2-Dichloropropane	7.069	77	152549	46.192	ug/l	98
27) cis-1,2-Dichloroethene	7.075	96	138303	52.038	ug/l	93
28) Bromochloromethane	7.422	49	90655	54.543	ug/l #	77
29) Tetrahydrofuran	7.440	42	90685	264.824	ug/l	87
30) Chloroform	7.593	83	215123	51.968	ug/l	98
31) Cyclohexane	7.875	56	154765	49.210	ug/l	90
32) 1,1,1-Trichloroethane	7.793	97	177230	51.216	ug/l	100
36) 1,1-Dichloropropene	8.005	75	147929	52.185	ug/l	95
37) Ethyl Acetate	7.163	43	60774	49.508	ug/l	99
38) Carbon Tetrachloride	7.987	117	151930	50.917	ug/l	98
39) Methylcyclohexane	9.275	83	179803	54.396	ug/l	93
40) Benzene	8.252	78	477148	52.577	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070824\
 Data File : VD079356.D
 Acq On : 08 Jul 2024 22:15
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024
 Supervised By :Semsettin Yesilyurt 07/09/2024

Quant Time: Jul 09 05:01:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070824S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 17:06:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.399	41	30387	43.023	ug/l	93
42) 1,2-Dichloroethane	8.322	62	112082	51.087	ug/l	98
43) Isopropyl Acetate	8.358	43	116201	51.799	ug/l	97
44) Trichloroethene	9.028	130	108767	50.277	ug/l	94
45) 1,2-Dichloropropane	9.305	63	118213	53.361	ug/l	98
46) Dibromomethane	9.393	93	70059	53.935	ug/l	91
47) Bromodichloromethane	9.581	83	164571	51.819	ug/l	100
48) Methyl methacrylate	9.375	41	55370	50.151	ug/l	89
49) 1,4-Dioxane	9.381	88	15501	1053.772	ug/l #	87
51) 4-Methyl-2-Pentanone	10.157	43	322757	268.703	ug/l	99
52) Toluene	10.334	92	305355	54.656	ug/l	99
53) t-1,3-Dichloropropene	10.552	75	150763	52.401	ug/l	99
54) cis-1,3-Dichloropropene	10.016	75	182319	53.050	ug/l	97
55) 1,1,2-Trichloroethane	10.728	97	94941	54.447	ug/l	93
56) Ethyl methacrylate	10.599	69	114792	48.747	ug/l	97
57) 1,3-Dichloropropane	10.875	76	151393	52.915	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.869	63	257415	248.874	ug/l	91
59) 2-Hexanone	10.916	43	225526	262.785	ug/l	99
60) Dibromochloromethane	11.075	129	118058	52.745	ug/l	95
61) 1,2-Dibromoethane	11.175	107	87912	53.470	ug/l	97
64) Tetrachloroethene	10.810	164	98379	55.506	ug/l	91
65) Chlorobenzene	11.604	112	329622	51.940	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.675	131	115044	53.730	ug/l	97
67) Ethyl Benzene	11.681	91	572030	53.993	ug/l	97
68) m/p-Xylenes	11.793	106	453258	110.794	ug/l	91
69) o-Xylene	12.116	106	205576	53.703	ug/l	85
70) Styrene	12.134	104	381455	55.843	ug/l	96
71) Bromoform	12.293	173	65594	52.877	ug/l #	96
73) Isopropylbenzene	12.416	105	533667	53.301	ug/l	98
74) N-amyl acetate	12.228	43	116987	51.992	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.669	83	114329	54.048	ug/l	93
76) 1,2,3-Trichloropropane	12.722	75	78692m	54.600	ug/l	
77) Bromobenzene	12.698	156	126727	53.491	ug/l	82
78) n-propylbenzene	12.757	91	678252	54.350	ug/l	95
79) 2-Chlorotoluene	12.846	91	387740	54.175	ug/l	91
80) 1,3,5-Trimethylbenzene	12.898	105	462572	54.507	ug/l	94
81) trans-1,4-Dichloro-2-b...	12.463	75	38116	49.863	ug/l	96
82) 4-Chlorotoluene	12.945	91	408774	53.668	ug/l	92
83) tert-Butylbenzene	13.163	119	378904	53.466	ug/l	91
84) 1,2,4-Trimethylbenzene	13.210	105	478837	55.624	ug/l	95
85) sec-Butylbenzene	13.340	105	589518	54.273	ug/l	97
86) p-Isopropyltoluene	13.457	119	488514	54.307	ug/l	96
87) 1,3-Dichlorobenzene	13.451	146	259191	51.879	ug/l	96
88) 1,4-Dichlorobenzene	13.534	146	257465	51.201	ug/l	96
89) n-Butylbenzene	13.787	91	469079	53.641	ug/l	97
90) Hexachloroethane	14.051	117	104253	52.460	ug/l	90
91) 1,2-Dichlorobenzene	13.828	146	228318	51.516	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.440	75	15360	51.697	ug/l	81
93) 1,2,4-Trichlorobenzene	15.092	180	124398	49.949	ug/l	95
94) Hexachlorobutadiene	15.198	225	68903	52.166	ug/l	94
95) Naphthalene	15.328	128	266153	53.756	ug/l	99
96) 1,2,3-Trichlorobenzene	15.516	180	115154	51.762	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070824\
Data File : VD079356.D
Acq On : 08 Jul 2024 22:15
Operator : RP/MD
Sample : VSTDCCC050
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024

Quant Time: Jul 09 05:01:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070824S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 17:06:59 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 17647
Ion Ratio Lower Upper
168 100
99 57.7 41.1 61.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#2
Dichlorodifluoromethane
Concen: 47.589 ug/l
RT: 1.928 min Scan# 58
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Tgt Ion: 85 Resp: 77865
Ion Ratio Lower Upper
85 100
87 33.1 16.0 48.0









#3
Chloromethane
Concen: 46.624 ug/l
RT: 2.146 min Scan# 95
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 50 Resp: 10214
Ion Ratio Lower Upper
50 100
52 33.6 28.2 42.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#4
Vinyl Chloride
Concen: 45.167 ug/l
RT: 2.275 min Scan# 117
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion: 62 Resp: 101698
Ion Ratio Lower Upper
62 100
64 33.9 26.3 39.5





#5
Bromomethane
Concen: 59.752 ug/l
RT: 2.681 min Scan# 11
Delta R.T. 0.023 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 94 Resp: 85249
Ion Ratio Lower Upper
94 100
96 94.1 76.9 115.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#6
Chloroethane
Concen: 63.808 ug/l
RT: 2.828 min Scan# 211
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Tgt Ion: 64 Resp: 76980
Ion Ratio Lower Upper
64 100
66 30.8 24.5 36.7





#7
Trichlorofluoromethane
Concen: 50.083 ug/l
RT: 3.164 min Scan# 20
Delta R.T. -0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion:101 Resp: 16477
Ion Ratio Lower Upper
101 100
103 63.9 53.2 79.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#8
Diethyl Ether
Concen: 49.975 ug/l
RT: 3.593 min Scan# 341
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Tgt Ion: 74 Resp: 52525
Ion Ratio Lower Upper
74 100
45 84.7 37.9 113.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.625 ug/l
RT: 3.958 min Scan# 403
Delta R.T. -0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 10447
Ion Ratio Lower Upper
101 100
85 41.8 33.5 50.3
151 71.9 60.3 90.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#10
Methyl Iodide
Concen: 54.143 ug/l
RT: 4.158 min Scan# 437
Delta R.T. -0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion:142 Resp: 123242
Ion Ratio Lower Upper
142 100
127 44.2 34.2 51.4
141 15.1 11.7 17.5





#11
Tert butyl alcohol
Concen: 218.860 ug/l m
RT: 5.064 min Scan# 591
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 59 Resp: 29054
Ion Ratio Lower Upper
59 100
57 11.1 3.6 5.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#12
1,1-Dichloroethene
Concen: 49.574 ug/l
RT: 3.934 min Scan# 399
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Tgt Ion: 96 Resp: 101596
Ion Ratio Lower Upper
96 100
61 134.9 107.3 160.9
98 62.6 51.7 77.5





#13
Acrolein
Concen: 175.788 ug/l
RT: 3.799 min Scan# 31
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 2893
Ion Ratio Lower Upper
56 100
55 72.2 55.3 82.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#14
Allyl chloride
Concen: 48.707 ug/l
RT: 4.558 min Scan# 505
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion: 41 Resp: 127203
Ion Ratio Lower Upper
41 100
39 60.3 58.0 87.0
76 49.1 45.9 68.9





#15
Acrylonitrile
Concen: 261.335 ug/l
RT: 5.252 min Scan# 61
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 12651
Ion Ratio Lower Upper
53 100
52 79.6 67.7 101.5
51 35.5 28.6 42.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#16
Acetone
Concen: 234.818 ug/l
RT: 4.022 min Scan# 414
Delta R.T. 0.005 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Tgt Ion: 43 Resp: 92877
Ion Ratio Lower Upper
43 100
58 38.2 30.4 45.6









#17
Carbon Disulfide
Concen: 49.794 ug/l
RT: 4.264 min Scan# 41
Delta R.T. -0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 76 Resp: 32239
Ion Ratio Lower Upper
76 100
78 9.9 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#18
Methyl Acetate
Concen: 50.673 ug/l
RT: 4.558 min Scan# 505
Delta R.T. -0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion: 43 Resp: 62712
Ion Ratio Lower Upper
43 100
74 28.8 28.1 42.1





#19
Methyl tert-butyl Ether
Concen: 52.711 ug/l
RT: 5.316 min Scan# 61
Delta R.T. 0.005 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 73 Resp: 240074
Ion Ratio Lower Upper
73 100
57 22.0 18.3 27.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#20
Methylene Chloride
Concen: 59.704 ug/l
RT: 4.793 min Scan# 545
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion: 84 Resp: 179679
Ion Ratio Lower Upper
84 100
49 98.0 75.6 113.4
51 33.1 24.2 36.2
86 61.5 49.2 73.8





#21
trans-1,2-Dichloroethene
Concen: 51.787 ug/l
RT: 5.311 min Scan# 611
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 11789
Ion Ratio Lower Upper
96 100
61 121.1 84.5 126.7
98 60.2 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#22
Diisopropyl ether
Concen: 53.514 ug/l
RT: 6.216 min Scan# 787
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Tgt Ion: 45 Resp: 312958
Ion Ratio Lower Upper
45 100
43 51.9 41.7 62.5
87 32.2 30.2 45.2
59 13.7 11.0 16.6









#23
Vinyl Acetate
Concen: 265.654 ug/l
RT: 6.152 min Scan# 776
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 1188930
Ion Ratio Lower Upper
43 100
86 13.3 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#24
1,1-Dichloroethane
Concen: 51.414 ug/l
RT: 6.111 min Scan# 769
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion: 63 Resp: 202734
Ion Ratio Lower Upper
63 100
98 7.6 4.7 14.0
100 4.3 2.9 8.6





#25
2-Butanone
Concen: 249.828 ug/l
RT: 7.081 min Scan# 91
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 140358
Ion Ratio Lower Upper
43 100
72 30.9 26.4 39.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#26
2,2-Dichloropropane
Concen: 46.192 ug/l
RT: 7.069 min Scan# 932
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Tgt Ion: 77 Resp: 152549
Ion Ratio Lower Upper
77 100
97 25.9 12.4 37.2





#27
cis-1,2-Dichloroethene
Concen: 52.038 ug/l
RT: 7.075 min Scan# 91
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 96 Resp: 13830
Ion Ratio Lower Upper
96 100
61 126.0 0.0 229.0
98 64.0 0.0 127.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#28
Bromochloromethane
Concen: 54.543 ug/l
RT: 7.422 min Scan# 992
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15



Tgt Ion: 49 Resp: 90655
Ion Ratio Lower Upper
49 100
129 2.6 0.0 7.0
130 84.7 87.7 131.5#





#29

Tetrahydrofuran

Concen: 264.824 ug/l

RT: 7.440 min Scan# 996

Delta R.T. 0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15

Instrument :

MSVOA_D

ClientSampleId :





VSTDCCC050EC

Tgt Ion:	42	Resp:	9068
Ion	Ratio	Lower	Upper
42	100		
72	52.0	49.4	74.0
71	47.1	45.0	67.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#30

Chloroform

Concen: 51.968 ug/l




RT: 7.593 min Scan# 1021


Delta R.T. 0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15

Tgt Ion:	83	Resp:	215123
Ion	Ratio	Lower	Upper
83	100		
85	66.6	51.8	77.6





#31

Cyclohexane

Concen: 49.210 ug/l

RT: 7.875 min Scan# 1069

Delta R.T. 0.000 min

Lab File: VD079356.D


Acq: 08 Jul 2024 22:15

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 56 Resp: 15476

Ion Ratio Lower Upper

56 100

69 34.9 28.0 42.0




84 96.0 87.4 131.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024

Supervised By :Semsettin Yesilyurt 07/09/2024



#32

1,1,1-Trichloroethane

Concen: 51.216 ug/l

RT: 7.793 min Scan# 1055

Delta R.T. 0.006 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15




Tgt Ion: 97 Resp: 177230


Ion Ratio Lower Upper

97 100

99 63.7 51.3 76.9

61 39.5 31.4 47.2





#33

1,2-Dichloroethane-d4

Concen: 50.853 ug/l

RT: 8.228 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 10575

Ion Ratio Lower Upper

65 100





67 55.9 0.0 111.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024

Supervised By :Semsettin Yesilyurt 07/09/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1222

Delta R.T. 0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15




Tgt Ion:114 Resp: 316296


Ion Ratio Lower Upper

114 100

63 18.7 0.0 34.0

88 14.5 0.0 27.2





#35

Dibromofluoromethane

Concen: 51.652 ug/l

RT: 7.805 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VD079356.D


Acq: 08 Jul 2024 22:15

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC






Tgt Ion:	113	Resp:	11525
Ion	Ratio	Lower	Upper
113	100		
111	103.6	84.0	126.0
192	17.8	15.4	23.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#36

1,1-Dichloropropene

Concen: 52.185 ug/l




RT: 8.005 min Scan# 1091


Delta R.T. 0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15

Tgt Ion:	75	Resp:	147929
Ion	Ratio	Lower	Upper
75	100		
110	35.1	19.6	58.7
77	30.5	26.1	39.1





#37

Ethyl Acetate

Concen: 49.508 ug/l

RT: 7.163 min Scan# 948

Delta R.T. 0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 60774

Ion	Ratio	Lower	Upper
43	100		
61	17.3	13.9	20.9
70	13.1	9.8	14.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#38

Carbon Tetrachloride

Concen: 50.917 ug/l

RT: 7.987 min Scan# 1088




Delta R.T. 0.000 min


Lab File: VD079356.D

Acq: 08 Jul 2024 22:15

Tgt Ion:117 Resp: 151930

Ion	Ratio	Lower	Upper
117	100		
119	92.0	75.4	113.0
121	29.5	25.0	37.4





#39

Methylcyclohexane

Concen: 54.396 ug/l

RT: 9.275 min Scan# 11307

Delta R.T. 0.006 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 17980

Ion Ratio Lower Upper

83 100

55 65.8 50.1 75.1





98 45.5 41.9 62.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024

Supervised By :Semsettin Yesilyurt 07/09/2024



#40

Benzene

Concen: 52.577 ug/l

RT: 8.252 min Scan# 1133

Delta R.T. 0.006 min

Lab File: VD079356.D




Acq: 08 Jul 2024 22:15


Tgt Ion: 78 Resp: 477148

Ion Ratio Lower Upper

78 100

77 25.2 18.1 27.1





#41
Methacrylonitrile
Concen: 43.023 ug/l
RT: 7.399 min Scan# 911
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Instrument :

MSVOA_D





ClientSampleId :

VSTDCCC050EC

Tgt Ion: 41 Resp: 3038
Ion Ratio Lower Upper
41 100
39 56.8 41.4 62.0
67 97.6 85.8 128.6
52 38.1 30.4 45.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#42
1,2-Dichloroethane
Concen: 51.087 ug/l
RT: 8.322 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion: 62 Resp: 112082
Ion Ratio Lower Upper
62 100
98 12.0 0.0 25.8





#43
Isopropyl Acetate
Concen: 51.799 ug/l
RT: 8.358 min Scan# 1151
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 11620
Ion Ratio Lower Upper
43 100
61 35.5 29.3 43.9
87 16.2 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#44
Trichloroethene
Concen: 50.277 ug/l
RT: 9.028 min Scan# 1265
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15



Tgt Ion:130 Resp: 108767
Ion Ratio Lower Upper
130 100
95 94.3 0.0 177.8





#45

1,2-Dichloropropane

Concen: 53.361 ug/l

RT: 9.305 min Scan# 11821

Delta R.T. -0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 11821

Ion Ratio Lower Upper

63 100





65 28.5 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024

Supervised By :Semsettin Yesilyurt 07/09/2024



#46

Dibromomethane

Concen: 53.935 ug/l

RT: 9.393 min Scan# 1327

Delta R.T. -0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15




Tgt Ion: 93 Resp: 70059


Ion Ratio Lower Upper

93 100

95 82.6 65.6 98.4


174 86.4 82.3 123.5





#47
Bromodichloromethane
Concen: 51.819 ug/l
RT: 9.581 min Scan# 1359
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 83 Resp: 16457
Ion Ratio Lower Upper
83 100
85 62.8 50.1 75.1
127 8.5 6.8 10.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#48
Methyl methacrylate
Concen: 50.151 ug/l
RT: 9.375 min Scan# 1324
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Tgt Ion: 41 Resp: 55370
Ion Ratio Lower Upper
41 100
69 107.6 96.7 145.1
39 49.3 43.8 65.6





#49
1,4-Dioxane
Concen: 1053.772 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 88 Resp: 1550
Ion Ratio Lower Upper
88 100
43 28.7 21.5 32.3
58 65.9 42.2 63.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#50
Toluene-d8
Concen: 52.821 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Tgt Ion: 98 Resp: 454344
Ion Ratio Lower Upper
98 100
100 62.4 51.6 77.4









#51
4-Methyl-2-Pentanone
Concen: 268.703 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 32275
Ion Ratio Lower Upper
43 100
58 44.8 36.6 55.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#52
Toluene
Concen: 54.656 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion: 92 Resp: 305355
Ion Ratio Lower Upper
92 100
91 178.2 141.7 212.5





#53
t-1,3-Dichloropropene
Concen: 52.401 ug/l
RT: 10.552 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 75 Resp: 15076
Ion Ratio Lower Upper
75 100
77 32.0 25.0 37.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#54
cis-1,3-Dichloropropene
Concen: 53.050 ug/l
RT: 10.016 min Scan# 1433
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Tgt Ion: 75 Resp: 182319
Ion Ratio Lower Upper
75 100
77 31.7 24.4 36.6
39 32.8 24.2 36.2





#55
1,1,2-Trichloroethane
Concen: 54.447 ug/l
RT: 10.728 min Scan# 1554
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion:	97	Resp:	9494
Ion Ratio	Lower	Upper	
97	100		
83	84.0	65.8	98.8
85	50.1	43.8	65.8
99	59.5	55.4	83.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#56
Ethyl methacrylate
Concen: 48.747 ug/l
RT: 10.599 min Scan# 1532
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Tgt Ion:	69	Resp:	114792
Ion Ratio	Lower	Upper	
69	100		
41	52.4	39.8	59.6
39	24.5	20.2	30.4





#57

1,3-Dichloropropane

Concen: 52.915 ug/l

RT: 10.875 min Scan# 1579

Delta R.T. 0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 15139

Ion Ratio Lower Upper

76 100





78 33.6 26.2 39.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024

Supervised By :Semsettin Yesilyurt 07/09/2024



#58

2-Chloroethyl Vinyl ether

Concen: 248.874 ug/l

RT: 9.869 min Scan# 1408

Delta R.T. 0.000 min

Lab File: VD079356.D




Acq: 08 Jul 2024 22:15


Tgt Ion: 63 Resp: 257415

Ion Ratio Lower Upper

63 100


106 27.8 26.4 39.6





#59
2-Hexanone
Concen: 262.785 ug/l
RT: 10.916 min Scan# 1586
Delta R.T. -0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 225520
Ion Ratio Lower Upper
43 100
58 63.0 31.1 93.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#60
Dibromochloromethane
Concen: 52.745 ug/l
RT: 11.075 min Scan# 1613
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion:129 Resp: 118058
Ion Ratio Lower Upper
129 100
127 79.8 38.0 113.9





#61
1,2-Dibromoethane
Concen: 53.470 ug/l
RT: 11.175 min Scan# 1630
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:107 Resp: 87912
Ion Ratio Lower Upper
107 100
109 93.2 72.0 108.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#62
4-Bromofluorobenzene
Concen: 54.480 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Tgt Ion: 95 Resp: 145651
Ion Ratio Lower Upper
95 100
174 76.3 0.0 195.4
176 73.6 0.0 192.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1699
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:117 Resp: 302348
Ion Ratio Lower Upper
117 100
82 54.4 39.8 59.6
119 32.3 25.7 38.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#64
Tetrachloroethene
Concen: 55.506 ug/l
RT: 10.810 min Scan# 1568
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Tgt Ion:164 Resp: 98379
Ion Ratio Lower Upper
164 100
166 119.6 108.6 163.0
129 95.8 74.6 112.0
131 93.5 68.6 103.0





#65
Chlorobenzene
Concen: 51.940 ug/l
RT: 11.604 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:112 Resp: 32962
Ion Ratio Lower Upper
112 100
114 32.2 26.0 39.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 53.730 ug/l
RT: 11.675 min Scan# 1715
Delta R.T. -0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15



Tgt Ion:131 Resp: 115044
Ion Ratio Lower Upper
131 100
133 95.5 49.2 147.6
119 68.4 33.4 100.2









#67
Ethyl Benzene
Concen: 53.993 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 572030
Ion Ratio Lower Upper
91 100
106 30.4 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#68
m/p-Xylenes
Concen: 110.794 ug/l
RT: 11.793 min Scan# 1735
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion:106 Resp: 453258
Ion Ratio Lower Upper
106 100
91 199.2 148.5 222.7





#69
o-Xylene
Concen: 53.703 ug/l
RT: 12.116 min Scan# 1790
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion:106 Resp: 205570
Ion Ratio Lower Upper
106 100
91 215.4 96.7 290.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#70
Styrene
Concen: 55.843 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion:104 Resp: 381455
Ion Ratio Lower Upper
104 100
78 46.6 33.0 49.6
103 55.2 44.2 66.4





#71
Bromoform
Concen: 52.877 ug/l
RT: 12.293 min Scan# 1820
Delta R.T. -0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:173 Resp: 65594
Ion Ratio Lower Upper
173 100
175 50.1 23.7 71.1
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Tgt Ion:152 Resp: 150004
Ion Ratio Lower Upper
152 100
115 61.5 26.5 79.5
150 174.3 0.0 341.6





#73
Isopropylbenzene
Concen: 53.301 ug/l
RT: 12.416 min Scan# 1841
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion:105 Resp: 53366
Ion Ratio Lower Upper
105 100
120 26.1 13.6 40.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#74
N-ethyl acetate
Concen: 51.992 ug/l
RT: 12.228 min Scan# 1809
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion: 43 Resp: 116987
Ion Ratio Lower Upper
43 100
70 50.7 44.3 66.5
55 27.4 23.2 34.8
61 28.0 24.2 36.4





#75
1,1,2,2-Tetrachloroethane
Concen: 54.048 ug/l
RT: 12.669 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 114329
Ion Ratio Lower Upper
83 100
131 9.5 5.5 16.5
85 61.3 33.8 101.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#76
1,2,3-Trichloropropane
Concen: 54.600 ug/l m
RT: 12.722 min Scan# 1893
Delta R.T. -0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion: 75 Resp: 78692
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 53.491 ug/l
RT: 12.698 min Scan# 1889
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:156 Resp: 12672
Ion Ratio Lower Upper
156 100
77 189.3 77.3 231.9
158 95.6 49.6 149.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#78
n-propylbenzene
Concen: 54.350 ug/l
RT: 12.757 min Scan# 1899
Delta R.T. -0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Tgt Ion: 91 Resp: 678252
Ion Ratio Lower Upper
91 100
120 22.0 12.3 36.8





#79
2-Chlorotoluene
Concen: 54.175 ug/l
RT: 12.846 min Scan# 1914
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 387740
Ion Ratio Lower Upper
91 100
126 33.3 19.3 57.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#80
1,3,5-Trimethylbenzene
Concen: 54.507 ug/l
RT: 12.898 min Scan# 1923
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Tgt Ion:105 Resp: 462572
Ion Ratio Lower Upper
105 100
120 47.4 25.7 77.1





#81
trans-1,4-Dichloro-2-butene
Concen: 49.863 ug/l
RT: 12.463 min Scan# 1849
Delta R.T. -0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 38110
Ion Ratio Lower Upper
75 100
53 71.6 61.4 92.2
89 45.3 35.4 53.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#82
4-Chlorotoluene
Concen: 53.668 ug/l
RT: 12.945 min Scan# 1931
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion: 91 Resp: 408774
Ion Ratio Lower Upper
91 100
126 33.4 18.9 56.9





#83
tert-Butylbenzene
Concen: 53.466 ug/l
RT: 13.163 min Scan# 1968
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:119 Resp: 378904
Ion Ratio Lower Upper
119 100
91 65.6 28.2 84.7
134 26.2 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024




#84
1,2,4-Trimethylbenzene
Concen: 55.624 ug/l
RT: 13.210 min Scan# 1976
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15



Tgt Ion:105 Resp: 478837
Ion Ratio Lower Upper
105 100
120 43.6 23.6 71.0





#85

sec-Butylbenzene

Concen: 54.273 ug/l

RT: 13.340 min Scan# 1998

Delta R.T. 0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion:105 Resp: 589518

Ion Ratio Lower Upper

105 100





134 19.1 10.3 30.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024

Supervised By :Semsettin Yesilyurt 07/09/2024



#86

p-Isopropyltoluene

Concen: 54.307 ug/l

RT: 13.457 min Scan# 2018

Delta R.T. 0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15




Tgt Ion:119 Resp: 488514


Ion Ratio Lower Upper

119 100

134 25.7 13.7 41.1


91 23.6 10.7 32.1





#87
1,3-Dichlorobenzene
Concen: 51.879 ug/l
RT: 13.451 min Scan# 2017
Delta R.T. -0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 25919
Ion Ratio Lower Upper
146 100
111 40.8 18.5 55.5
148 64.0 30.9 92.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#88
1,4-Dichlorobenzene
Concen: 51.201 ug/l
RT: 13.534 min Scan# 2031
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion:146 Resp: 257465
Ion Ratio Lower Upper
146 100
111 39.7 17.5 52.6
148 63.9 31.4 94.2





#89
n-Butylbenzene
Concen: 53.641 ug/l
RT: 13.787 min Scan# 2074
Delta R.T. 0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 469079
Ion Ratio Lower Upper
91 100
92 52.5 25.7 77.0
134 24.2 13.6 40.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#90
Hexachloroethane
Concen: 52.460 ug/l
RT: 14.051 min Scan# 2119
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15


Tgt Ion:117 Resp: 104253
Ion Ratio Lower Upper
117 100
201 64.4 36.3 108.9





#91
1,2-Dichlorobenzene
Concen: 51.516 ug/l
RT: 13.828 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 228318
Ion Ratio Lower Upper
146 100
111 43.0 19.1 57.1
148 63.6 31.1 93.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.697 ug/l
RT: 14.440 min Scan# 2185
Delta R.T. -0.006 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15

Tgt Ion: 75 Resp: 15360
Ion Ratio Lower Upper
75 100
155 80.1 47.7 143.1
157 102.3 63.4 190.2





#93

1,2,4-Trichlorobenzene

Concen: 49.949 ug/l

RT: 15.092 min Scan# 21

Delta R.T. 0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion:180 Resp: 124398

Ion Ratio Lower Upper

180 100

182 96.4 46.9 140.7





145 33.5 14.0 41.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024

Supervised By :Semsettin Yesilyurt 07/09/2024



#94

Hexachlorobutadiene

Concen: 52.166 ug/l

RT: 15.198 min Scan# 2314

Delta R.T. 0.000 min

Lab File: VD079356.D

Acq: 08 Jul 2024 22:15




Tgt Ion:225 Resp: 68903


Ion Ratio Lower Upper

225 100

223 61.0 33.3 99.9


227 63.5 33.6 100.6





#95
Naphthalene
Concen: 53.756 ug/l
RT: 15.328 min Scan# 2336
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:128 Resp: 26615
Ion Ratio Lower Upper
128 100
127 13.4 10.2 15.4
129 10.4 8.5 12.7

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/09/2024
Supervised By :Semsettin Yesilyurt 07/09/2024



#96
1,2,3-Trichlorobenzene
Concen: 51.762 ug/l
RT: 15.516 min Scan# 2368
Delta R.T. 0.000 min
Lab File: VD079356.D
Acq: 08 Jul 2024 22:15



Tgt Ion:180 Resp: 115154
Ion Ratio Lower Upper
180 100
182 95.4 47.8 143.3
145 34.7 16.1 48.2

