

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070924\
 Data File : VD079359.D
 Acq On : 09 Jul 2024 13:53
 Operator : RP/MD
 Sample : VD0709SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0709SBL01

Quant Time: Jul 09 16:53:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070824S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 17:06:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	146753	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	291945	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	275225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	110803	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	99777	57.695	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.380%
35) Dibromofluoromethane	7.804	113	111212	53.997	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.000%
50) Toluene-d8	10.269	98	394891	49.738	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.480%
62) 4-Bromofluorobenzene	12.569	95	110300	44.698	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070924\
Data File : VD079359.D
Acq On : 09 Jul 2024 13:53
Operator : RP/MD
Sample : VD0709SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0709SBL01

Quant Time: Jul 09 16:53:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070824S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 17:06:59 2024
Response via : Initial Calibration

