

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072221\
 Data File : VD069808.D
 Acq On : 22 Jul 2021 15:23
 Operator : VA/SY
 Sample : VD0722SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0722SBL01

Quant Time: Jul 23 05:00:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	589127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1111608	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1016485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	418732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	366230	54.052	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.100%	
35) Dibromofluoromethane	7.914	113	352465	50.298	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.600%	
50) Toluene-d8	10.338	98	1293011	47.852	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 95.700%	
62) 4-Bromofluorobenzene	12.626	95	399358	44.224	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 88.440%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.962	84	105660	6.410	ug/l	93
95) Naphthalene	15.390	128	5323	5.349	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072221\
Data File : VD069808.D
Acq On : 22 Jul 2021 15:23
Operator : VA/SY
Sample : VD0722SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0722SBL01

Quant Time: Jul 23 05:00:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

