

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072221\
 Data File : VD069815.D
 Acq On : 22 Jul 2021 18:40
 Operator : VA/SY
 Sample : M3117-05
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NJ-CLEAN-5

Quant Time: Jul 23 05:03:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	10247	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	18130	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	18187	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	6649	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	10811	91.735	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 183.480%#	
35) Dibromofluoromethane	7.914	113	5829	51.001	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.000%	
50) Toluene-d8	10.338	98	17672	40.100	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 80.200%	
62) 4-Bromofluorobenzene	12.626	95	5980	40.602	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 81.200%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.967	84	10526	70.907	ug/l	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072221\
Data File : VD069815.D
Acq On : 22 Jul 2021 18:40
Operator : VA/SY
Sample : M3117-05
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NJ-CLEAN-5

Quant Time: Jul 23 05:03:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

