

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072624\
 Data File : VD079412.D
 Acq On : 26 Jul 2024 17:15
 Operator : RP/MD
 Sample : P3343-03
 Misc : 6.92G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-A-GRAB

Quant Time: Jul 27 03:35:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 26 12:23:49 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	60639	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	144628	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	76883	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	12900	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	63057	111.044	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 222.080%#	
35) Dibromofluoromethane	7.799	113	68193	78.199	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 156.400%#	
50) Toluene-d8	10.269	98	166952	48.431	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 96.860%	
62) 4-Bromofluorobenzene	12.569	95	29317	27.457	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 54.920%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.793	84	13305	6.065	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072624\
Data File : VD079412.D
Acq On : 26 Jul 2024 17:15
Operator : RP/MD
Sample : P3343-03
Misc : 6.92G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-A-GRAB

Quant Time: Jul 27 03:35:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072524S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 26 12:23:49 2024
Response via : Initial Calibration

