

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072624\
 Data File : VD079418.D
 Acq On : 26 Jul 2024 19:59
 Operator : RP/MD
 Sample : P3358-01
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-02-07242024

Quant Time: Jul 27 03:38:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 26 12:23:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	6380	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.769	114	13837	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	11293	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	4085	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	6931	116.126	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	232.260%#
35) Dibromofluoromethane	7.799	113	7214	86.734	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	173.460%#
50) Toluene-d8	10.269	98	13327	40.334	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	80.660%
62) 4-Bromofluorobenzene	12.569	95	4187	40.693	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072624\
Data File : VD079418.D
Acq On : 26 Jul 2024 19:59
Operator : RP/MD
Sample : P3358-01
Misc : 5.01G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-02-07242024

Quant Time: Jul 27 03:38:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072524S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 26 12:23:49 2024
Response via : Initial Calibration

