

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072924\
 Data File : VD079444.D
 Acq On : 29 Jul 2024 19:02
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Jul 30 03:07:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 02:34:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	99448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	357023	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	390353	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	179078	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	84870	84.005	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	168.000%#
35) Dibromofluoromethane	7.805	113	145649	60.273	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	120.540%
50) Toluene-d8	10.269	98	468402	49.914	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.820%
62) 4-Bromofluorobenzene	12.569	95	145045	53.412	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072924\
Data File : VD079444.D
Acq On : 29 Jul 2024 19:02
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Jul 30 03:07:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 02:34:50 2024
Response via : Initial Calibration

