

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073024\
 Data File : VD079460.D
 Acq On : 30 Jul 2024 15:53
 Operator : RP/MD
 Sample : P3341-01RE
 Misc : 5.10G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 AU-05-072324RE

Quant Time: Jul 31 02:51:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	15749	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	31795	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	34634	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	16306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	15205	95.034	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 190.060%#	
35) Dibromofluoromethane	7.804	113	15654	72.741	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 145.480%	
50) Toluene-d8	10.269	98	35694	42.711	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 85.420%	
62) 4-Bromofluorobenzene	12.569	95	12690	52.473	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 104.940%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.793	84	12131	53.682	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073024\
Data File : VD079460.D
Acq On : 30 Jul 2024 15:53
Operator : RP/MD
Sample : P3341-01RE
Misc : 5.10G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
AU-05-072324RE

Quant Time: Jul 31 02:51:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

