

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073124\
 Data File : VD079500.D
 Acq On : 31 Jul 2024 17:04
 Operator : RP/MD
 Sample : P3379-11RE
 Misc : 6.03G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OU4-VSL-01-072324RE

Quant Time: Aug 01 07:17:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	49840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	82491	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	88981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	44146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	54092	106.832	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 213.660%#	
35) Dibromofluoromethane	7.805	113	41282	73.937	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 147.880%#	
50) Toluene-d8	10.269	98	109094	50.315	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.620%	
62) 4-Bromofluorobenzene	12.569	95	35522	56.614	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 113.220%	
Target Compounds						Qvalue
20) Methylene Chloride	4.793	84	22175	23.088	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073124\
Data File : VD079500.D
Acq On : 31 Jul 2024 17:04
Operator : RP/MD
Sample : P3379-11RE
Misc : 6.03G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OU4-VSL-01-072324RE

Quant Time: Aug 01 07:17:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

