

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073124\
 Data File : VD079504.D
 Acq On : 31 Jul 2024 18:55
 Operator : RP/MD
 Sample : P3387-04RE
 Misc : 4.54G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-10-VOCRE

Quant Time: Aug 01 01:54:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	136988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	230107	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	155087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	33362	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	94612	67.985	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.960%
35) Dibromofluoromethane	7.804	113	100607	64.596	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	129.200%
50) Toluene-d8	10.269	98	267665	44.255	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.520%
62) 4-Bromofluorobenzene	12.569	95	42927	24.526	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	49.060%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.793	84	44962	10.709	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073124\
Data File : VD079504.D
Acq On : 31 Jul 2024 18:55
Operator : RP/MD
Sample : P3387-04RE
Misc : 4.54G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-10-VOCRE

Quant Time: Aug 01 01:54:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

