

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073124\
 Data File : VD079510.D
 Acq On : 31 Jul 2024 23:30
 Operator : RP/MD
 Sample : VD0731SBS02
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0731SBS02

Quant Time: Aug 01 01:58:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 08/01/2024
 Supervised By :Mahesh Dadoda 08/01/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	199079	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	340667	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	326827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	168486	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	118250	58.469	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 116.940%			
35) Dibromofluoromethane	7.805	113	123981	53.769	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 107.540%			
50) Toluene-d8	10.269	98	482982	53.939	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 107.880%			
62) 4-Bromofluorobenzene	12.569	95	142956	55.170	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 110.340%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.928	85	30923	19.825	ug/l	98
3) Chloromethane	2.146	50	60576	21.077	ug/l	94
4) Vinyl Chloride	2.281	62	73145	21.404	ug/l	100
5) Bromomethane	2.675	94	59782	26.477	ug/l	97
6) Chloroethane	2.828	64	45106	20.541	ug/l	97
7) Trichlorofluoromethane	3.170	101	70929	20.871	ug/l	99
8) Diethyl Ether	3.587	74	22144	21.571	ug/l	80
9) 1,1,2-Trichlorotrifluo...	3.958	101	44637	20.767	ug/l	97
10) Methyl Iodide	4.158	142	51244	21.160	ug/l	97
11) Tert butyl alcohol	5.064	59	12501	136.004	ug/l #	91
12) 1,1-Dichloroethene	3.940	96	43615	20.276	ug/l	93
13) Acrolein	3.793	56	17718	119.195	ug/l	97
14) Allyl chloride	4.558	41	57952	20.403	ug/l	87
15) Acrylonitrile	5.252	53	53325	111.161	ug/l	98
16) Acetone	4.022	43	36684	93.332	ug/l	99
17) Carbon Disulfide	4.264	76	150618	21.213	ug/l	98
18) Methyl Acetate	4.564	43	33392	27.363	ug/l #	87
19) Methyl tert-butyl Ether	5.316	73	96743	22.034	ug/l	92
20) Methylene Chloride	4.799	84	91421	24.544	ug/l #	86
21) trans-1,2-Dichloroethene	5.305	96	51157	21.171	ug/l	88
22) Diisopropyl ether	6.216	45	139487	22.071	ug/l #	95
23) Vinyl Acetate	6.152	43	491196	103.393	ug/l	95
24) 1,1-Dichloroethane	6.105	63	89692	21.178	ug/l	99
25) 2-Butanone	7.087	43	57254	102.747	ug/l	99
26) 2,2-Dichloropropane	7.075	77	65828	19.216	ug/l	96
27) cis-1,2-Dichloroethene	7.081	96	58210	21.075	ug/l	87
28) Bromochloromethane	7.428	49	39977	21.526	ug/l #	73
29) Tetrahydrofuran	7.446	42	38374	111.122	ug/l #	85
30) Chloroform	7.593	83	99037	23.092	ug/l	99
31) Cyclohexane	7.875	56	63757	18.982	ug/l #	88
32) 1,1,1-Trichloroethane	7.793	97	76141	21.515	ug/l	99
36) 1,1-Dichloropropene	8.010	75	63769	20.262	ug/l	96
37) Ethyl Acetate	7.169	43	29725	23.036	ug/l	97
38) Carbon Tetrachloride	7.993	117	70051	21.174	ug/l	96
39) Methylcyclohexane	9.275	83	68399	18.215	ug/l	91
40) Benzene	8.252	78	208092	20.818	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073124\
 Data File : VD079510.D
 Acq On : 31 Jul 2024 23:30
 Operator : RP/MD
 Sample : VD0731SBS02
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0731SBS02

Quant Time: Aug 01 01:58:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Romaben Patel 08/01/2024
 Supervised By :Mahesh Dadoda 08/01/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.399	41	15144	21.337	ug/l #	80
42) 1,2-Dichloroethane	8.328	62	53509	22.464	ug/l	99
43) Isopropyl Acetate	8.363	43	52479	21.640	ug/l #	93
44) Trichloroethene	9.028	130	51198	20.575	ug/l	95
45) 1,2-Dichloropropane	9.305	63	52905	21.748	ug/l	99
46) Dibromomethane	9.393	93	31232	23.159	ug/l	97
47) Bromodichloromethane	9.581	83	73368	22.157	ug/l #	97
48) Methyl methacrylate	9.381	41	21277	19.605	ug/l	90
49) 1,4-Dioxane	9.387	88	5880	420.791	ug/l #	75
51) 4-Methyl-2-Pentanone	10.157	43	140402	113.398	ug/l	92
52) Toluene	10.334	92	133069	21.525	ug/l	99
53) t-1,3-Dichloropropene	10.552	75	63591	20.796	ug/l	96
54) cis-1,3-Dichloropropene	10.016	75	77274	21.009	ug/l	96
55) 1,1,2-Trichloroethane	10.728	97	39851	22.433	ug/l #	91
56) Ethyl methacrylate	10.593	69	46345	22.018	ug/l	90
57) 1,3-Dichloropropane	10.875	76	67957	22.512	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.869	63	92018	94.794	ug/l	93
59) 2-Hexanone	10.922	43	95403	110.484	ug/l	92
60) Dibromochloromethane	11.069	129	51381	22.141	ug/l	95
61) 1,2-Dibromoethane	11.175	107	38378	23.060	ug/l	97
64) Tetrachloroethene	10.810	164	45745	20.757	ug/l	96
65) Chlorobenzene	11.604	112	149600	20.671	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.675	131	50349	21.145	ug/l	98
67) Ethyl Benzene	11.681	91	233049	19.670	ug/l	99
68) m/p-Xylenes	11.793	106	192643	40.984	ug/l	96
69) o-Xylene	12.116	106	84513	19.616	ug/l	89
70) Styrene	12.134	104	158637	21.002	ug/l	99
71) Bromoform	12.298	173	31907	23.703	ug/l #	100
73) Isopropylbenzene	12.416	105	215620	19.475	ug/l	99
74) N-amyl acetate	12.228	43	49300	19.894	ug/l	91
75) 1,1,2,2-Tetrachloroethane	12.669	83	48527	22.171	ug/l	94
76) 1,2,3-Trichloropropane	12.722	75	35818m	24.321	ug/l	
77) Bromobenzene	12.698	156	58495	20.705	ug/l	92
78) n-propylbenzene	12.757	91	276602	19.640	ug/l	97
79) 2-Chlorotoluene	12.846	91	155171	19.787	ug/l	97
80) 1,3,5-Trimethylbenzene	12.898	105	188283	20.061	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.463	75	15406	19.952	ug/l	97
82) 4-Chlorotoluene	12.946	91	165612	20.258	ug/l	100
83) tert-Butylbenzene	13.163	119	156071	19.550	ug/l	96
84) 1,2,4-Trimethylbenzene	13.204	105	187566	19.974	ug/l	99
85) sec-Butylbenzene	13.340	105	240124	19.235	ug/l	100
86) p-Isopropyltoluene	13.457	119	203434	19.769	ug/l	99
87) 1,3-Dichlorobenzene	13.451	146	121157	20.703	ug/l	99
88) 1,4-Dichlorobenzene	13.534	146	120687	20.977	ug/l	98
89) n-Butylbenzene	13.787	91	181780	19.186	ug/l	99
90) Hexachloroethane	14.045	117	44088	20.192	ug/l	94
91) 1,2-Dichlorobenzene	13.828	146	104905	20.975	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.440	75	6592	22.634	ug/l	91
93) 1,2,4-Trichlorobenzene	15.092	180	57846	18.357	ug/l	97
94) Hexachlorobutadiene	15.198	225	32187	18.925	ug/l	97
95) Naphthalene	15.328	128	107766	19.759	ug/l	97
96) 1,2,3-Trichlorobenzene	15.516	180	51810	18.691	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073124\
Data File : VD079510.D
Acq On : 31 Jul 2024 23:30
Operator : RP/MD
Sample : VD0731SBS02
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02


Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024

Quant Time: Aug 01 01:58:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion:168 Resp: 199079
Ion Ratio Lower Upper
168 100
99 50.1 41.1 61.7




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#2
Dichlorodifluoromethane
Concen: 19.825 ug/l
RT: 1.928 min Scan# 58
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Tgt Ion: 85 Resp: 30923
Ion Ratio Lower Upper
85 100
87 30.9 16.0 48.0









#3
Chloromethane
Concen: 21.077 ug/l
RT: 2.146 min Scan# 95
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02

Tgt Ion: 50 Resp: 60570
Ion Ratio Lower Upper
50 100
52 31.7 28.2 42.2




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#4
Vinyl Chloride
Concen: 21.404 ug/l
RT: 2.281 min Scan# 118
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion: 62 Resp: 73145
Ion Ratio Lower Upper
62 100
64 32.7 26.3 39.5





#5
Bromomethane
Concen: 26.477 ug/l
RT: 2.675 min Scan# 185
Delta R.T. -0.012 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion: 94 Resp: 5978
Ion Ratio Lower Upper
94 100
96 92.8 76.9 115.3




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#6
Chloroethane
Concen: 20.541 ug/l
RT: 2.828 min Scan# 211
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion: 64 Resp: 45106
Ion Ratio Lower Upper
64 100
66 32.2 24.5 36.7





#7
Trichlorofluoromethane
Concen: 20.871 ug/l
RT: 3.170 min Scan# 20
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion:101 Resp: 70929
Ion Ratio Lower Upper
101 100
103 65.5 53.2 79.8




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#8
Diethyl Ether
Concen: 21.571 ug/l
RT: 3.587 min Scan# 340
Delta R.T. -0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Tgt Ion: 74 Resp: 22144
Ion Ratio Lower Upper
74 100
45 92.9 37.9 113.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.767 ug/l
RT: 3.958 min Scan# 403
Delta R.T. -0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02

Tgt Ion:101 Resp: 4463
Ion Ratio Lower Upper
101 100
85 45.8 33.5 50.3
151 73.8 60.3 90.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#10
Methyl Iodide
Concen: 21.160 ug/l
RT: 4.158 min Scan# 437
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion:142 Resp: 51244
Ion Ratio Lower Upper
142 100
127 44.3 34.2 51.4
141 16.2 11.7 17.5





#11
Tert butyl alcohol
Concen: 136.004 ug/l
RT: 5.064 min Scan# 591
Delta R.T. 0.018 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02




Tgt Ion: 59 Resp: 1250
Ion Ratio Lower Upper
59 100
57 7.6 3.6 5.4

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024




#12
1,1-Dichloroethene
Concen: 20.276 ug/l
RT: 3.940 min Scan# 400
Delta R.T. 0.012 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Tgt Ion: 96 Resp: 43615
Ion Ratio Lower Upper
96 100
61 125.4 107.3 160.9
98 60.6 51.7 77.5





#13
Acrolein
Concen: 119.195 ug/l
RT: 3.793 min Scan# 31
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion: 56 Resp: 17718
Ion Ratio Lower Upper
56 100
55 71.8 55.3 82.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#14
Allyl chloride
Concen: 20.403 ug/l
RT: 4.558 min Scan# 505
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion: 41 Resp: 57952
Ion Ratio Lower Upper
41 100
39 62.8 58.0 87.0
76 46.3 45.9 68.9





#15
Acrylonitrile
Concen: 111.161 ug/l
RT: 5.252 min Scan# 61
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion: 53 Resp: 5332
Ion Ratio Lower Upper
53 100
52 84.2 67.7 101.5
51 38.7 28.6 42.8




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#16
Acetone
Concen: 93.332 ug/l
RT: 4.022 min Scan# 414
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion: 43 Resp: 36684
Ion Ratio Lower Upper
43 100
58 37.1 30.4 45.6





#17
Carbon Disulfide
Concen: 21.213 ug/l
RT: 4.264 min Scan# 455
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion: 76 Resp: 150618
Ion Ratio Lower Upper
76 100
78 9.4 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#18
Methyl Acetate
Concen: 27.363 ug/l
RT: 4.564 min Scan# 506
Delta R.T. -0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Tgt Ion: 43 Resp: 33392
Ion Ratio Lower Upper
43 100
74 27.4 28.1 42.1#









#19
Methyl tert-butyl Ether
Concen: 22.034 ug/l
RT: 5.316 min Scan# 611
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02

Tgt Ion: 73 Resp: 9674
Ion Ratio Lower Upper
73 100
57 26.6 18.3 27.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#20
Methylene Chloride
Concen: 24.544 ug/l
RT: 4.799 min Scan# 546
Delta R.T. -0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Tgt Ion: 84 Resp: 91421
Ion Ratio Lower Upper
84 100
49 113.5 75.6 113.4#
51 38.3 24.2 36.2#
86 65.5 49.2 73.8









#21
trans-1,2-Dichloroethene
Concen: 21.171 ug/l
RT: 5.305 min Scan# 611
Delta R.T. -0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02

Tgt Ion: 96 Resp: 5115
Ion Ratio Lower Upper
96 100
61 122.8 84.5 126.7
98 60.8 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#22
Diisopropyl ether
Concen: 22.071 ug/l
RT: 6.216 min Scan# 787
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion: 45 Resp: 139487
Ion Ratio Lower Upper
45 100
43 52.3 41.7 62.5
87 29.7 30.2 45.2#
59 14.5 11.0 16.6





#23
 Vinyl Acetate
 Concen: 103.393 ug/l
 RT: 6.152 min Scan# 776
 Delta R.T. 0.000 min
 Lab File: VD079510.D
 Acq: 31 Jul 2024 23:30




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0731SBS02




Tgt Ion: 43 Resp: 491190
 Ion Ratio Lower Upper
 43 100
 86 13.2 12.3 18.5

Manual Integrations
 APPROVED



Reviewed By :Romaben Patel 08/01/2024
 Supervised By :Mahesh Dadoda 08/01/2024



#24
 1,1-Dichloroethane
 Concen: 21.178 ug/l
 RT: 6.105 min Scan# 768
 Delta R.T. 0.000 min
 Lab File: VD079510.D
 Acq: 31 Jul 2024 23:30



Tgt Ion: 63 Resp: 89692
 Ion Ratio Lower Upper
 63 100
 98 8.9 4.7 14.0
 100 5.9 2.9 8.6





#25

2-Butanone

Concen: 102.747 ug/l

RT: 7.087 min Scan# 91

Delta R.T. -0.006 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Instrument :

MSVOA_D

ClientSampleId :

VD0731SBS02

Tgt Ion: 43 Resp: 57254

Ion Ratio Lower Upper

43 100





72 33.4 26.4 39.6

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/01/2024

Supervised By :Mahesh Dadoda 08/01/2024



#26

2,2-Dichloropropane

Concen: 19.216 ug/l

RT: 7.075 min Scan# 933

Delta R.T. -0.006 min

Lab File: VD079510.D




Acq: 31 Jul 2024 23:30


Tgt Ion: 77 Resp: 65828

Ion Ratio Lower Upper

77 100

97 26.6 12.4 37.2









#27
cis-1,2-Dichloroethene
Concen: 21.075 ug/l
RT: 7.081 min Scan# 91
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02

Tgt Ion: 96 Resp: 58210
Ion Ratio Lower Upper
96 100
61 133.0 0.0 229.0
98 68.4 0.0 127.2




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#28
Bromochloromethane
Concen: 21.526 ug/l
RT: 7.428 min Scan# 993
Delta R.T. 0.012 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Tgt Ion: 49 Resp: 39977
Ion Ratio Lower Upper
49 100
129 1.9 0.0 7.0
130 80.8 87.7 131.5#





#29

Tetrahydrofuran

Concen: 111.122 ug/l

RT: 7.446 min Scan# 996

Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Instrument :

MSVOA_D

ClientSampleId :





VD0731SBS02

Tgt Ion:	42	Resp:	38374
Ion Ratio	Lower	Upper	
42	100		
72	49.3	49.4	74.0
71	46.4	45.0	67.6

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#30

Chloroform

Concen: 23.092 ug/l




RT: 7.593 min Scan# 1021


Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30


Tgt Ion:	83	Resp:	99037
Ion Ratio	Lower	Upper	
83	100		
85	65.4	51.8	77.6





#31
Cyclohexane
Concen: 18.982 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. -0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02




Tgt Ion: 56 Resp: 6375
Ion Ratio Lower Upper
56 100
69 45.1 28.0 42.0
84 97.8 87.4 131.0

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024




#32
1,1,1-Trichloroethane
Concen: 21.515 ug/l
RT: 7.793 min Scan# 1055
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30



Tgt Ion: 97 Resp: 76141
Ion Ratio Lower Upper
97 100
99 64.5 51.3 76.9
61 40.9 31.4 47.2





#33

1,2-Dichloroethane-d4

Concen: 58.469 ug/l

RT: 8.228 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Instrument :

MSVOA_D

ClientSampleId :

VD0731SBS02

Tgt Ion: 65 Resp: 118250

Ion Ratio Lower Upper

65 100





67 54.1 0.0 111.0

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/01/2024

Supervised By :Mahesh Dadoda 08/01/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1222

Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30




Tgt Ion:114 Resp: 340667

Ion Ratio Lower Upper

114 100

63 17.9 0.0 34.0

88 15.0 0.0 27.2





#35

Dibromofluoromethane

Concen: 53.769 ug/l

RT: 7.805 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Instrument :

MSVOA_D

ClientSampleId :





VD0731SBS02

Tgt Ion:	113	Resp:	12398
Ion Ratio	Lower	Upper	
113	100		
111	101.9	84.0	126.0
192	20.0	15.4	23.0

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#36

1,1-Dichloropropene

Concen: 20.262 ug/l




RT: 8.010 min Scan# 1092


Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Tgt Ion:	75	Resp:	63769
Ion Ratio	Lower	Upper	
75	100		
110	36.6	19.6	58.7
77	30.6	26.1	39.1





#37

Ethyl Acetate

Concen: 23.036 ug/l

RT: 7.169 min Scan# 948

Delta R.T. -0.006 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Instrument :

MSVOA_D

ClientSampleId :

VD0731SBS02





Tgt Ion: 43 Resp: 29725

Ion	Ratio	Lower	Upper
43	100		
61	16.5	13.9	20.9
70	10.1	9.8	14.6

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#38

Carbon Tetrachloride

Concen: 21.174 ug/l

RT: 7.993 min Scan# 1089




Delta R.T. 0.006 min


Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Tgt Ion:117 Resp: 70051

Ion	Ratio	Lower	Upper
117	100		
119	97.1	75.4	113.0
121	34.6	25.0	37.4





#39

Methylcyclohexane

Concen: 18.215 ug/l

RT: 9.275 min Scan# 11307

Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Instrument :

MSVOA_D

ClientSampleId :

VD0731SBS02





Tgt Ion: 83	Resp: 68399
Ion Ratio	Lower Upper
83 100	
55 72.7	50.1 75.1
98 50.0	41.9 62.9

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/01/2024

Supervised By :Mahesh Dadoda 08/01/2024



#40

Benzene

Concen: 20.818 ug/l




RT: 8.252 min Scan# 1133


Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30


Tgt Ion: 78	Resp: 208092
Ion Ratio	Lower Upper
78 100	
77 24.2	18.1 27.1





#41
Methacrylonitrile
Concen: 21.337 ug/l
RT: 7.399 min Scan# 91
Delta R.T. -0.018 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion: 41 Resp: 1514
Ion Ratio Lower Upper
41 100
39 29.4 41.4 62.0
67 84.5 85.8 128.6
52 37.7 30.4 45.6




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#42
1,2-Dichloroethane
Concen: 22.464 ug/l
RT: 8.328 min Scan# 1146
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Tgt Ion: 62 Resp: 53509
Ion Ratio Lower Upper
62 100
98 12.7 0.0 25.8









#43
Isopropyl Acetate
Concen: 21.640 ug/l
RT: 8.363 min Scan# 1152
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02

Tgt Ion: 43 Resp: 52479
Ion Ratio Lower Upper
43 100
61 33.6 29.3 43.9
87 14.3 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#44
Trichloroethene
Concen: 20.575 ug/l
RT: 9.028 min Scan# 1265
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Tgt Ion:130 Resp: 51198
Ion Ratio Lower Upper
130 100
95 93.2 0.0 177.8





#45

1,2-Dichloropropane

Concen: 21.748 ug/l

RT: 9.305 min Scan# 1312

Delta R.T. 0.006 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Instrument :

MSVOA_D

ClientSampleId :

VD0731SBS02

Tgt Ion: 63 Resp: 5290

Ion Ratio Lower Upper

63 100





65 30.5 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/01/2024

Supervised By :Mahesh Dadoda 08/01/2024



#46

Dibromomethane

Concen: 23.159 ug/l

RT: 9.393 min Scan# 1327

Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30




Tgt Ion: 93 Resp: 31232


Ion Ratio Lower Upper

93 100

95 81.5 65.6 98.4


174 97.8 82.3 123.5





#47
Bromodichloromethane
Concen: 22.157 ug/l
RT: 9.581 min Scan# 1359
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion: 83 Resp: 7336
Ion Ratio Lower Upper
83 100
85 65.0 50.1 75.1
127 10.2 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#48
Methyl methacrylate
Concen: 19.605 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Tgt Ion: 41 Resp: 21277
Ion Ratio Lower Upper
41 100
69 110.1 96.7 145.1
39 63.2 43.8 65.6





#49

1,4-Dioxane

Concen: 420.791 ug/l

RT: 9.387 min Scan# 1325

Delta R.T. -0.006 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Instrument :

MSVOA_D

ClientSampleId :

VD0731SBS02

Tgt Ion: 88 Resp: 5880

Ion Ratio Lower Upper

88 100

43 35.5 21.5 32.3





58 73.5 42.2 63.4

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/01/2024

Supervised By :Mahesh Dadoda 08/01/2024



#50

Toluene-d8

Concen: 53.939 ug/l

RT: 10.269 min Scan# 1476

Delta R.T. 0.006 min

Lab File: VD079510.D




Acq: 31 Jul 2024 23:30


Tgt Ion: 98 Resp: 482982

Ion Ratio Lower Upper

98 100


100 63.8 51.6 77.4





#51
4-Methyl-2-Pentanone
Concen: 113.398 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion: 43 Resp: 140402
Ion Ratio Lower Upper
43 100
58 40.7 36.6 55.0




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#52
Toluene
Concen: 21.525 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion: 92 Resp: 133069
Ion Ratio Lower Upper
92 100
91 176.3 141.7 212.5





#53
t-1,3-Dichloropropene
Concen: 20.796 ug/l
RT: 10.552 min Scan# 11
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion: 75 Resp: 6359
Ion Ratio Lower Upper
75 100
77 33.8 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#54
cis-1,3-Dichloropropene
Concen: 21.009 ug/l
RT: 10.016 min Scan# 1433
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion: 75 Resp: 77274
Ion Ratio Lower Upper
75 100
77 30.5 24.4 36.6
39 35.0 24.2 36.2





#55
1,1,2-Trichloroethane
Concen: 22.433 ug/l
RT: 10.728 min Scan# 1554
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02




Tgt Ion	Ratio	Lower	Upper
97	100		
83	86.8	65.8	98.8
85	67.4	43.8	65.8
99	64.1	55.4	83.0

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024




#56
Ethyl methacrylate
Concen: 22.018 ug/l
RT: 10.593 min Scan# 1531
Delta R.T. -0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30



Tgt Ion	Ratio	Lower	Upper
69	100		
41	58.0	39.8	59.6
39	28.1	20.2	30.4





#57

1,3-Dichloropropane

Concen: 22.512 ug/l

RT: 10.875 min Scan# 1579

Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Instrument :

MSVOA_D

ClientSampleId :

VD0731SBS02

Tgt Ion: 76 Resp: 6795

Ion Ratio Lower Upper

76 100





78 33.1 26.2 39.2

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/01/2024

Supervised By :Mahesh Dadoda 08/01/2024



#58

2-Chloroethyl Vinyl ether

Concen: 94.794 ug/l

RT: 9.869 min Scan# 1408

Delta R.T. 0.000 min

Lab File: VD079510.D




Acq: 31 Jul 2024 23:30


Tgt Ion: 63 Resp: 92018

Ion Ratio Lower Upper

63 100

106 28.8 26.4 39.6









#59
2-Hexanone
Concen: 110.484 ug/l
RT: 10.922 min Scan# 110
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02

Tgt Ion: 43 Resp: 9540
Ion Ratio Lower Upper
43 100
58 56.4 31.1 93.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#60
Dibromochloromethane
Concen: 22.141 ug/l
RT: 11.069 min Scan# 1612
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion:129 Resp: 51381
Ion Ratio Lower Upper
129 100
127 80.4 38.0 113.9





#61
1,2-Dibromoethane
Concen: 23.060 ug/l
RT: 11.175 min Scan# 1630
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02




Tgt Ion:107 Resp: 38378
Ion Ratio Lower Upper
107 100
109 93.2 72.0 108.0

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#62
4-Bromofluorobenzene
Concen: 55.170 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Tgt Ion: 95 Resp: 142956
Ion Ratio Lower Upper
95 100
174 91.3 0.0 195.4
176 91.8 0.0 192.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1699
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion:117 Resp: 32682
Ion Ratio Lower Upper
117 100
82 52.9 39.8 59.6
119 32.2 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#64
Tetrachloroethene
Concen: 20.757 ug/l
RT: 10.810 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion:164 Resp: 45745
Ion Ratio Lower Upper
164 100
166 126.7 108.6 163.0
129 94.7 74.6 112.0
131 86.4 68.6 103.0





#65
Chlorobenzene
Concen: 20.671 ug/l
RT: 11.604 min Scan# 1703
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02




Tgt Ion:112 Resp: 149600
Ion Ratio Lower Upper
112 100
114 29.6 26.0 39.0

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 21.145 ug/l
RT: 11.675 min Scan# 1715
Delta R.T. -0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Tgt Ion:131 Resp: 50349
Ion Ratio Lower Upper
131 100
133 100.9 49.2 147.6
119 67.5 33.4 100.2





#67
Ethyl Benzene
Concen: 19.670 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02




Tgt Ion: 91 Resp: 233049
Ion Ratio Lower Upper
91 100
106 31.7 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024




#68
m/p-Xylenes
Concen: 40.984 ug/l
RT: 11.793 min Scan# 1735
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30



Tgt Ion:106 Resp: 192643
Ion Ratio Lower Upper
106 100
91 190.9 148.5 222.7









#69
o-Xylene
Concen: 19.616 ug/l
RT: 12.116 min Scan# 1790
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02

Tgt Ion:106 Resp: 8451
Ion Ratio Lower Upper
106 100
91 209.4 96.7 290.1




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#70
Styrene
Concen: 21.002 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion:104 Resp: 158637
Ion Ratio Lower Upper
104 100
78 43.0 33.0 49.6
103 55.7 44.2 66.4





#71
Bromoform
Concen: 23.703 ug/l
RT: 12.298 min Scan# 1821
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion:173 Resp: 3190
Ion Ratio Lower Upper
173 100
175 47.5 23.7 71.1
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion:152 Resp: 168486
Ion Ratio Lower Upper
152 100
115 55.6 26.5 79.5
150 166.0 0.0 341.6





#73
Isopropylbenzene
Concen: 19.475 ug/l
RT: 12.416 min Scan# 1841
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02




Tgt Ion:105 Resp: 215620
Ion Ratio Lower Upper
105 100
120 27.4 13.6 40.7

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024




#74
N-ethyl acetate
Concen: 19.894 ug/l
RT: 12.228 min Scan# 1809
Delta R.T. -0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Tgt Ion: 43 Resp: 49300
Ion Ratio Lower Upper
43 100
70 47.6 44.3 66.5
55 25.7 23.2 34.8
61 26.7 24.2 36.4





#75
1,1,2,2-Tetrachloroethane
Concen: 22.171 ug/l
RT: 12.669 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02




Tgt Ion: 83 Resp: 4852
Ion Ratio Lower Upper
83 100
131 10.2 5.5 16.5
85 62.6 33.8 101.3

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024




#76
1,2,3-Trichloropropane
Concen: 24.321 ug/l m
RT: 12.722 min Scan# 1893
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30



Tgt Ion: 75 Resp: 35818
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 20.705 ug/l
RT: 12.698 min Scan# 1889
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02

Tgt Ion:156 Resp: 5849
Ion Ratio Lower Upper
156 100
77 169.7 77.3 231.9
158 97.7 49.6 149.0




Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#78
n-propylbenzene
Concen: 19.640 ug/l
RT: 12.757 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Tgt Ion: 91 Resp: 276602
Ion Ratio Lower Upper
91 100
120 23.0 12.3 36.8









#79
2-Chlorotoluene
Concen: 19.787 ug/l
RT: 12.846 min Scan# 1914
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02

Tgt Ion: 91 Resp: 15517
Ion Ratio Lower Upper
91 100
126 36.7 19.3 57.8




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#80
1,3,5-Trimethylbenzene
Concen: 20.061 ug/l
RT: 12.898 min Scan# 1923
Delta R.T. -0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion:105 Resp: 188283
Ion Ratio Lower Upper
105 100
120 50.6 25.7 77.1





#81
trans-1,4-Dichloro-2-butene
Concen: 19.952 ug/l
RT: 12.463 min Scan# 1849
Delta R.T. -0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion: 75 Resp: 15400
Ion Ratio Lower Upper
75 100
53 79.7 61.4 92.2
89 43.5 35.4 53.0




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#82
4-Chlorotoluene
Concen: 20.258 ug/l
RT: 12.946 min Scan# 1931
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion: 91 Resp: 165612
Ion Ratio Lower Upper
91 100
126 37.7 18.9 56.9





#83
tert-Butylbenzene
Concen: 19.550 ug/l
RT: 13.163 min Scan# 1968
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02




Tgt Ion:119 Resp: 15607
Ion Ratio Lower Upper
119 100
91 60.7 28.2 84.7
134 25.7 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024




#84
1,2,4-Trimethylbenzene
Concen: 19.974 ug/l
RT: 13.204 min Scan# 1975
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30



Tgt Ion:105 Resp: 187566
Ion Ratio Lower Upper
105 100
120 47.8 23.6 71.0





#85

sec-Butylbenzene

Concen: 19.235 ug/l

RT: 13.340 min Scan# 1998

Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Instrument :

MSVOA_D

ClientSampleId :

VD0731SBS02

Tgt Ion:105 Resp: 240124

Ion Ratio Lower Upper

105 100





134 20.6 10.3 30.8

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/01/2024

Supervised By :Mahesh Dadoda 08/01/2024



#86

p-Isopropyltoluene

Concen: 19.769 ug/l

RT: 13.457 min Scan# 2018

Delta R.T. 0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30




Tgt Ion:119 Resp: 203434


Ion Ratio Lower Upper

119 100

134 26.7 13.7 41.1

91 21.7 10.7 32.1









#87
 1,3-Dichlorobenzene
 Concen: 20.703 ug/l
 RT: 13.451 min Scan# 2017
 Delta R.T. 0.000 min
 Lab File: VD079510.D
 Acq: 31 Jul 2024 23:30

Instrument : MSVOA_D
 ClientSampleId : VD0731SBS02

Tgt Ion:146 Resp: 12115

Ion	Ratio	Lower	Upper
146	100		
111	37.4	18.5	55.5
148	62.4	30.9	92.8




Manual Integrations
APPROVED
 Reviewed By :Romaben Patel 08/01/2024
 Supervised By :Mahesh Dadoda 08/01/2024




#88
 1,4-Dichlorobenzene
 Concen: 20.977 ug/l
 RT: 13.534 min Scan# 2031
 Delta R.T. 0.000 min
 Lab File: VD079510.D
 Acq: 31 Jul 2024 23:30

Tgt Ion:146 Resp: 120687


Ion	Ratio	Lower	Upper
146	100		
111	37.6	17.5	52.6
148	63.1	31.4	94.2





#89
n-Butylbenzene
Concen: 19.186 ug/l
RT: 13.787 min Scan# 2074
Delta R.T. 0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02



Tgt Ion: 91 Resp: 181780
Ion Ratio Lower Upper
91 100
92 50.6 25.7 77.0
134 26.9 13.6 40.8




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#90
Hexachloroethane
Concen: 20.192 ug/l
RT: 14.045 min Scan# 2118
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Tgt Ion:117 Resp: 44088
Ion Ratio Lower Upper
117 100
201 77.4 36.3 108.9









#91
1,2-Dichlorobenzene
Concen: 20.975 ug/l
RT: 13.828 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30

Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02

Tgt Ion:146 Resp: 10490
Ion Ratio Lower Upper
146 100
111 39.7 19.1 57.1
148 63.6 31.1 93.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 22.634 ug/l
RT: 14.440 min Scan# 2185
Delta R.T. -0.006 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30


Tgt Ion: 75 Resp: 6592
Ion Ratio Lower Upper
75 100
155 90.5 47.7 143.1
157 112.9 63.4 190.2





#93
1,2,4-Trichlorobenzene
Concen: 18.357 ug/l
RT: 15.092 min Scan# 2297
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30




Instrument :
MSVOA_D
ClientSampleId :
VD0731SBS02




Tgt Ion:180 Resp: 57840
Ion Ratio Lower Upper
180 100
182 96.2 46.9 140.7
145 29.7 14.0 41.9

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/01/2024
Supervised By :Mahesh Dadoda 08/01/2024




#94
Hexachlorobutadiene
Concen: 18.925 ug/l
RT: 15.198 min Scan# 2314
Delta R.T. 0.000 min
Lab File: VD079510.D
Acq: 31 Jul 2024 23:30



Tgt Ion:225 Resp: 32187
Ion Ratio Lower Upper
225 100
223 67.1 33.3 99.9
227 62.8 33.6 100.6





#95

Naphthalene

Concen: 19.759 ug/l

RT: 15.328 min Scan# 2336

Delta R.T. -0.000 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Instrument :

MSVOA_D

ClientSampleId :

VD0731SBS02

Tgt Ion:128 Resp: 107760

Ion Ratio Lower Upper

128 100

127 14.2 10.2 15.4





129 11.9 8.5 12.7

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/01/2024

Supervised By :Mahesh Dadoda 08/01/2024



#96

1,2,3-Trichlorobenzene

Concen: 18.691 ug/l

RT: 15.516 min Scan# 2368

Delta R.T. -0.006 min

Lab File: VD079510.D

Acq: 31 Jul 2024 23:30

Tgt Ion:180 Resp: 51810

Ion Ratio Lower Upper

180 100

182 99.9 47.8 143.3

145 32.1 16.1 48.2

