

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080124\
 Data File : VD079533.D
 Acq On : 01 Aug 2024 18:45
 Operator : RP/MD
 Sample : P3374-10RE
 Misc : 7.54G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-862-K2-SO-12.75-13.25-072024RE

Quant Time: Aug 02 01:32:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	13195	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	29974	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	43750	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	22294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	39257	292.856	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	585.720%#
35) Dibromofluoromethane	7.804	113	22270	109.770	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	219.540%#
50) Toluene-d8	10.269	98	41612	52.817	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	105.640%
62) 4-Bromofluorobenzene	12.569	95	19078	83.680	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	167.360%#
Target Compounds						
					Qvalue	
16) Acetone	4.022	43	2258	86.675	ug/l	97
20) Methylene Chloride	4.805	84	8578	42.944	ug/l	93
27) cis-1,2-Dichloroethene	7.081	96	2638	14.410	ug/l	97
44) Trichloroethene	9.028	130	5638	25.752	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080124\
Data File : VD079533.D
Acq On : 01 Aug 2024 18:45
Operator : RP/MD
Sample : P3374-10RE
Misc : 7.54G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-862-K2-SO-12.75-13.25-072024RE

Quant Time: Aug 02 01:32:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

