

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080124\
 Data File : VD079534.D
 Acq On : 01 Aug 2024 19:13
 Operator : RP/MD
 Sample : P3374-13RE
 Misc : 6.51G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-847-G1-SO-2.5-3.0-072024RE

Quant Time: Aug 02 01:32:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	25710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	50587	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	56853	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	25783	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	40287	154.245	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 308.480%#	
35) Dibromofluoromethane	7.805	113	27700	80.900	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 161.800%#	
50) Toluene-d8	10.269	98	59338	44.627	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 89.260%	
62) 4-Bromofluorobenzene	12.569	95	21413	55.651	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 111.300%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.799	84	7018	4.505	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080124\
Data File : VD079534.D
Acq On : 01 Aug 2024 19:13
Operator : RP/MD
Sample : P3374-13RE
Misc : 6.51G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-847-G1-SO-2.5-3.0-072024RE

Quant Time: Aug 02 01:32:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

