

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080524\
 Data File : VD079580.D
 Acq On : 05 Aug 2024 14:59
 Operator : RP/MD
 Sample : P3463-03
 Misc : 5.05G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-05-082024

Quant Time: Aug 05 16:06:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	203444	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	390533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	355198	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	157014	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	117337	56.772	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.540%
35) Dibromofluoromethane	7.805	113	544	0.206	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.420%#
50) Toluene-d8	10.269	98	446658	43.513	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	87.020%
62) 4-Bromofluorobenzene	12.569	95	121707	40.972	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080524\
Data File : VD079580.D
Acq On : 05 Aug 2024 14:59
Operator : RP/MD
Sample : P3463-03
Misc : 5.05G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-05-082024

Quant Time: Aug 05 16:06:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

