

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080524\
 Data File : VD079606.D
 Acq On : 06 Aug 2024 04:39
 Operator : RP/MD
 Sample : P3374-06RE
 Misc : 8.35G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-861-K2-SO-13.5-14-072024RE

Quant Time: Aug 06 05:14:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	132075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	259696	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	244374	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	95338	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	73764	54.976	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.960%	
35) Dibromofluoromethane	7.804	113	84357	47.992	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 95.980%	
50) Toluene-d8	10.269	98	287928	42.181	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 84.360%	
62) 4-Bromofluorobenzene	12.569	95	72917	36.914	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 73.820%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	3.964	101	23756	16.659	ug/l	96
12) 1,1-Dichloroethene	3.946	96	3483	2.441	ug/l #	74
16) Acetone	4.028	43	5931	22.745	ug/l	96
27) cis-1,2-Dichloroethene	7.081	96	42987	23.459	ug/l	96
44) Trichloroethene	9.028	130	1521182	801.935	ug/l	94
64) Tetrachloroethene	10.810	164	14868	9.023	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080524\
Data File : VD079606.D
Acq On : 06 Aug 2024 04:39
Operator : RP/MD
Sample : P3374-06RE
Misc : 8.35G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-861-K2-SO-13.5-14-072024RE

Quant Time: Aug 06 05:14:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

