

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080524\
 Data File : VD079611.D
 Acq On : 06 Aug 2024 06:55
 Operator : RP/MD
 Sample : P3374-19MS
 Misc : 7.36G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-863-K2-SO-12.75-13.25-072024MS

Quant Time: Aug 06 08:10:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Romaben Patel 08/06/2024
 Supervised By :Mahesh Dadoda 08/08/2024

Supervised By :Mahesh
 Dadoda
 08/08/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	63152	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	155140	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	110308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	27108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	79744	124.296	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 248.600%#			
35) Dibromofluoromethane	7.805	113	79289	75.509	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 151.020%#			
50) Toluene-d8	10.269	98	206317	50.596	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 101.200%			
62) 4-Bromofluorobenzene	12.569	95	32664	27.681	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 55.360%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.934	85	39365	79.555	ug/l	96
3) Chloromethane	2.146	50	99150	108.751	ug/l	97
4) Vinyl Chloride	2.281	62	115497	106.543	ug/l	97
5) Bromomethane	2.693	94	72630	109.272	ug/l	98
6) Chloroethane	2.834	64	75189	107.939	ug/l	92
7) Trichlorofluoromethane	3.175	101	84832	78.689	ug/l	99
8) Diethyl Ether	3.599	74	37277	114.468	ug/l	86
9) 1,1,2-Trichlorotrifluo...	3.964	101	46437	68.105	ug/l	99
10) Methyl Iodide	4.164	142	63334	72.782	ug/l	97
11) Tert butyl alcohol	5.052	59	16523	505.892	ug/l #	90
12) 1,1-Dichloroethene	3.940	96	82356	120.695	ug/l	93
13) Acrolein	3.787	56	53	Below Cal	#	15
14) Allyl chloride	4.564	41	35980	39.932	ug/l	91
15) Acrylonitrile	5.252	53	81143	533.225	ug/l	96
16) Acetone	4.022	43	102110	818.958	ug/l	91
17) Carbon Disulfide	4.269	76	204296	90.704	ug/l	100
18) Methyl Acetate	4.564	43	75575	195.227	ug/l #	85
19) Methyl tert-butyl Ether	5.316	73	163133	117.125	ug/l	96
20) Methylene Chloride	4.799	84	141197	155.750	ug/l #	87
21) trans-1,2-Dichloroethene	5.311	96	181100	236.257	ug/l	87
22) Diisopropyl ether	6.211	45	218944	109.211	ug/l #	93
23) Vinyl Acetate	6.164	43	120049	79.659	ug/l #	89
24) 1,1-Dichloroethane	6.111	63	216178	160.908	ug/l	96
25) 2-Butanone	7.087	43	146769	830.304	ug/l #	1
26) 2,2-Dichloropropane	7.075	77	90927	83.672	ug/l #	1
27) cis-1,2-Dichloroethene	7.075	96	19365871	22102.707	ug/l	98
28) Bromochloromethane	7.428	49	60481	102.661	ug/l #	68
29) Tetrahydrofuran	7.446	42	61309	559.661	ug/l #	81
30) Chloroform	7.593	83	145013	106.586	ug/l	99
31) Cyclohexane	7.875	56	49930	46.860	ug/l #	82
32) 1,1,1-Trichloroethane	7.793	97	97367	86.729	ug/l	96
36) 1,1-Dichloropropene	8.010	75	66486	46.389	ug/l	96
37) Ethyl Acetate	7.169	43	38604	65.695	ug/l #	64
38) Carbon Tetrachloride	7.993	117	62655	41.586	ug/l	94
39) Methylcyclohexane	9.275	83	38358	22.431	ug/l	93
40) Benzene	8.252	78	312269	68.598	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080524\
 Data File : VD079611.D
 Acq On : 06 Aug 2024 06:55
 Operator : RP/MD
 Sample : P3374-19MS
 Misc : 7.36G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-863-K2-SO-12.75-13.25-072024MS

Quant Time: Aug 06 08:10:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Romaben Patel 08/06/2024
 Supervised By :Mahesh Dadoda 08/08/2024

Supervised By :Mahesh
 Dadoda
 08/08/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
41) Methacrylonitrile	7.393	41	22833	70.641	ug/l	#	86
42) 1,2-Dichloroethane	8.322	62	87942	81.071	ug/l		96
43) Isopropyl Acetate	8.363	43	64920	58.783	ug/l	#	82
44) Trichloroethene	9.028	130	497177	438.743	ug/l		90
45) 1,2-Dichloropropane	9.305	63	80764	72.905	ug/l		96
46) Dibromomethane	9.393	93	44785	72.924	ug/l		91
47) Bromodichloromethane	9.587	83	91247	60.510	ug/l		98
48) Methyl methacrylate	9.381	41	34799	70.408	ug/l		85
49) 1,4-Dioxane	9.387	88	9482	1490.032	ug/l		92
51) 4-Methyl-2-Pentanone	10.157	43	198113	351.360	ug/l		93
52) Toluene	10.334	92	151956	53.974	ug/l		99
53) t-1,3-Dichloropropene	10.552	75	61531	44.185	ug/l		96
54) cis-1,3-Dichloropropene	10.016	75	74680	44.583	ug/l		96
55) 1,1,2-Trichloroethane	10.728	97	56711	70.100	ug/l		95
56) Ethyl methacrylate	10.599	69	56431	58.870	ug/l	#	88
57) 1,3-Dichloropropane	10.875	76	93096	67.720	ug/l		100
58) 2-Chloroethyl Vinyl ether	9.875	63	13979	31.622	ug/l		92
59) 2-Hexanone	10.922	43	103705	263.720	ug/l		96
60) Dibromochloromethane	11.069	129	58221	55.090	ug/l		99
61) 1,2-Dibromoethane	11.181	107	47576	62.773	ug/l		96
64) Tetrachloroethene	10.810	164	60455	81.278	ug/l		94
65) Chlorobenzene	11.604	112	146545	59.995	ug/l		96
66) 1,1,1,2-Tetrachloroethane	11.681	131	54096	67.312	ug/l		98
67) Ethyl Benzene	11.681	91	206826	51.723	ug/l		99
68) m/p-Xylenes	11.793	106	159305	100.416	ug/l		91
69) o-Xylene	12.116	106	79858	54.918	ug/l		88
70) Styrene	12.134	104	66892	26.239	ug/l	#	85
71) Bromoform	12.293	173	29708	65.389	ug/l	#	97
73) Isopropylbenzene	12.416	105	151905	73.019	ug/l		98
74) N-amyl acetate	12.234	43	9954	24.965	ug/l	#	89
75) 1,1,2,2-Tetrachloroethane	12.669	83	57105	162.157	ug/l		94
76) 1,2,3-Trichloropropane	12.722	75	32309m	136.353	ug/l		
77) Bromobenzene	12.698	156	45585	100.288	ug/l		86
78) n-propylbenzene	12.757	91	156712	69.159	ug/l		98
79) 2-Chlorotoluene	12.845	91	105126	83.319	ug/l		96
80) 1,3,5-Trimethylbenzene	12.898	105	117543	77.841	ug/l		97
81) trans-1,4-Dichloro-2-b...	12.463	75	7969	64.144	ug/l		97
82) 4-Chlorotoluene	12.945	91	102226	77.719	ug/l		98
83) tert-Butylbenzene	13.163	119	92543	62.307	ug/l		95
84) 1,2,4-Trimethylbenzene	13.210	105	115914	66.979	ug/l		100
85) sec-Butylbenzene	13.340	105	104988	52.272	ug/l		99
86) p-Isopropyltoluene	13.457	119	86017	45.626	ug/l		99
87) 1,3-Dichlorobenzene	13.457	146	62097	65.952	ug/l		97
88) 1,4-Dichlorobenzene	13.534	146	61026	65.926	ug/l		98
89) n-Butylbenzene	13.787	91	53446	31.932	ug/l		98
90) Hexachloroethane	14.051	117	18737	53.337	ug/l		98
91) 1,2-Dichlorobenzene	13.828	146	58832	73.112	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	14.440	75	5960	127.193	ug/l		89
93) 1,2,4-Trichlorobenzene	15.098	180	14387	28.377	ug/l		97
94) Hexachlorobutadiene	15.198	225	4635	16.939	ug/l		94
95) Naphthalene	15.322	128	43986	50.126	ug/l		96
96) 1,2,3-Trichlorobenzene	15.516	180	13159	29.506	ug/l		97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080524\
Data File : VD079611.D
Acq On : 06 Aug 2024 06:55
Operator : RP/MD
Sample : P3374-19MS
Misc : 7.36G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 40 Sample Multiplier: 1

Quant Time: Aug 06 08:10:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

Instrument :
MSVOA_D
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024MS

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024

Supervised By :Mahesh
Dadoda
08/08/2024

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)
----------	------	---	-----	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080524\
Data File : VD079611.D
Acq On : 06 Aug 2024 06:55
Operator : RP/MD
Sample : P3374-19MS
Misc : 7.36G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 40 Sample Multiplier: 1


Instrument :
MSVOA_D
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024MS


Quant Time: Aug 06 08:10:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55



Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS




Tgt Ion:168 Resp: 63152
Ion Ratio Lower Upper
168 100
99 54.5 41.1 61.7

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024







Supervised By :Mahesh
Dadoda
08/08/2024



#2
Dichlorodifluoromethane
Concen: 79.555 ug/l
RT: 1.934 min Scan# 59
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 85 Resp: 39365
Ion Ratio Lower Upper
85 100
87 29.9 16.0 48.0





#3
Chloromethane
Concen: 108.751 ug/l
RT: 2.146 min Scan# 95
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55



Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS




Tgt Ion: 50 Resp: 99150
Ion Ratio Lower Upper
50 100
52 33.6 28.2 42.2

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024







Supervised By :Mahesh
Dadoda
08/08/2024



#4
Vinyl Chloride
Concen: 106.543 ug/l
RT: 2.281 min Scan# 118
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 62 Resp: 115497
Ion Ratio Lower Upper
62 100
64 31.3 26.3 39.5





#5
Bromomethane
Concen: 109.272 ug/l
RT: 2.693 min Scan# 188
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument :
MSVOA_D
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 94 Resp: 72630
Ion Ratio Lower Upper
94 100
96 97.8 76.9 115.3

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#6
Chloroethane
Concen: 107.939 ug/l
RT: 2.834 min Scan# 212
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 64 Resp: 75189
Ion Ratio Lower Upper
64 100
66 34.9 24.5 36.7





#7
 Trichlorofluoromethane
 Concen: 78.689 ug/l
 RT: 3.175 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: VD079611.D
 Acq: 06 Aug 2024 06:55

Instrument :
 MSVOA_D
 ClientSampleId :
 S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion:101 Resp: 8483
 Ion Ratio Lower Upper
 101 100
 103 66.0 53.2 79.8

Manual Integrations APPROVED




Reviewed By :Romaben Patel 08/06/2024
 Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
 Dadoda
 08/08/2024



#8
 Diethyl Ether
 Concen: 114.468 ug/l
 RT: 3.599 min Scan# 342
 Delta R.T. 0.006 min
 Lab File: VD079611.D
 Acq: 06 Aug 2024 06:55


Tgt Ion: 74 Resp: 37277
 Ion Ratio Lower Upper
 74 100
 45 87.4 37.9 113.6





#9
1,1,2-Trichlorotrifluoroethane
Concen: 68.105 ug/l
RT: 3.964 min Scan# 403
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument :
MSVOA_D
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion:101 Resp: 4643
Ion Ratio Lower Upper
101 100
85 42.6 33.5 50.3
151 74.6 60.3 90.5

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#10
Methyl Iodide
Concen: 72.782 ug/l
RT: 4.164 min Scan# 438
Delta R.T. 0.012 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion:142 Resp: 63334
Ion Ratio Lower Upper
142 100
127 45.1 34.2 51.4
141 14.6 11.7 17.5





#11
Tert butyl alcohol
Concen: 505.892 ug/l
RT: 5.052 min Scan# 51
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55



Instrument :
MSVOA_D
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024MS




Tgt Ion: 59 Resp: 1652
Ion Ratio Lower Upper
59 100
57 7.9 3.6 5.4

Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024





Supervised By :Mahesh
Dadoda
08/08/2024




#12
1,1-Dichloroethene
Concen: 120.695 ug/l
RT: 3.940 min Scan# 400
Delta R.T. 0.012 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55



Tgt Ion: 96 Resp: 82356
Ion Ratio Lower Upper
96 100
61 144.5 107.3 160.9
98 62.5 51.7 77.5





#14
 Allyl chloride
 Concen: 39.932 ug/l
 RT: 4.564 min Scan# 505
 Delta R.T. 0.012 min
 Lab File: VD079611.D
 Acq: 06 Aug 2024 06:55





Instrument :
 MSVOA_D
 ClientSampleId :
 S-863-K2-SO-12.75-13.25-072024MS

Tgt Ion: 41 Resp: 35980
 Ion Ratio Lower Upper
 41 100
 39 66.3 58.0 87.0
 76 48.4 45.9 68.9

Manual Integrations
 APPROVED




Reviewed By :Romaben Patel 08/06/2024
 Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
 Dadoda
 08/08/2024



#15
 Acrylonitrile
 Concen: 533.225 ug/l
 RT: 5.252 min Scan# 623
 Delta R.T. 0.000 min
 Lab File: VD079611.D
 Acq: 06 Aug 2024 06:55


Tgt Ion: 53 Resp: 81143
 Ion Ratio Lower Upper
 53 100
 52 81.3 67.7 101.5
 51 38.1 28.6 42.8





#16
Acetone
Concen: 818.958 ug/l
RT: 4.022 min Scan# 414
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55



Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS




Tgt Ion: 43 Resp: 102110
Ion Ratio Lower Upper
43 100
58 32.8 30.4 45.6

Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024





Supervised By :Mahesh
Dadoda
08/08/2024




#17
Carbon Disulfide
Concen: 90.704 ug/l
RT: 4.269 min Scan# 456
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55




Tgt Ion: 76 Resp: 204296
Ion Ratio Lower Upper
76 100
78 8.5 6.9 10.3





#18
Methyl Acetate
Concen: 195.227 ug/l
RT: 4.564 min Scan# 506
Delta R.T. -0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55



Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS




Tgt Ion: 43 Resp: 7557
Ion Ratio Lower Upper
43 100
74 26.7 28.1 42.1

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024







Supervised By :Mahesh
Dadoda
08/08/2024



#19
Methyl tert-butyl Ether
Concen: 117.125 ug/l
RT: 5.316 min Scan# 634
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 73 Resp: 163133
Ion Ratio Lower Upper
73 100
57 21.1 18.3 27.5





#20
Methylene Chloride
Concen: 155.750 ug/l
RT: 4.799 min Scan# 546
Delta R.T. -0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 84	Resp: 14119
Ion Ratio	Lower Upper
84 100	
49 114.2	75.6 113.4
51 35.6	24.2 36.2
86 63.7	49.2 73.8

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#21
trans-1,2-Dichloroethene
Concen: 236.257 ug/l
RT: 5.311 min Scan# 633
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 96	Resp: 181100
Ion Ratio	Lower Upper
96 100	
61 126.0	84.5 126.7
98 64.6	50.5 75.7





#22
Diisopropyl ether
Concen: 109.211 ug/l
RT: 6.211 min Scan# 786
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 45 Resp: 218944
Ion Ratio Lower Upper
45 100
43 53.9 41.7 62.5
87 29.9 30.2 45.2
59 11.8 11.0 16.6

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#23
Vinyl Acetate
Concen: 79.659 ug/l
RT: 6.164 min Scan# 778
Delta R.T. 0.012 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 43 Resp: 120049
Ion Ratio Lower Upper
43 100
86 10.9 12.3 18.5#





#24
1,1-Dichloroethane
Concen: 160.908 ug/l
RT: 6.111 min Scan# 70
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument :
MSVOA_D
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 63 Resp: 216178
Ion Ratio Lower Upper
63 100
98 7.5 4.7 14.0
100 5.0 2.9 8.6

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#25
2-Butanone
Concen: 830.304 ug/l
RT: 7.087 min Scan# 935
Delta R.T. -0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 43 Resp: 146769
Ion Ratio Lower Upper
43 100
72 122.4 26.4 39.6#





#26
2,2-Dichloropropane
Concen: 83.672 ug/l
RT: 7.075 min Scan# 91
Delta R.T. -0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument :
MSVOA_D
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024MS





Tgt Ion: 77 Resp: 9092
Ion Ratio Lower Upper
77 100
97 886.8 12.4 37.2

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#27
cis-1,2-Dichloroethene
Concen: 22102.707 ug/l
RT: 7.075 min Scan# 933
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 96 Resp:19365871
Ion Ratio Lower Upper
96 100
61 114.0 0.0 229.0
98 67.3 0.0 127.2





#28
Bromochloromethane
Concen: 102.661 ug/l
RT: 7.428 min Scan# 992
Delta R.T. 0.012 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument :
MSVOA_D
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 49 Resp: 6048
Ion Ratio Lower Upper
49 100
129 2.1 0.0 7.0
130 75.1 87.7 131.5

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#29
Tetrahydrofuran
Concen: 559.661 ug/l
RT: 7.446 min Scan# 996
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 42 Resp: 61309
Ion Ratio Lower Upper
42 100
72 46.1 49.4 74.0#
71 43.3 45.0 67.6#





#30
Chloroform
Concen: 106.586 ug/l
RT: 7.593 min Scan# 1021
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55



Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS




Tgt Ion: 83 Resp: 14501
Ion Ratio Lower Upper
83 100
85 65.8 51.8 77.6

Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024





Supervised By :Mahesh
Dadoda
08/08/2024




#31
Cyclohexane
Concen: 46.860 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. -0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55



Tgt Ion: 56 Resp: 49930
Ion Ratio Lower Upper
56 100
69 33.1 28.0 42.0
84 85.6 87.4 131.0#





#32

1,1,1-Trichloroethane

Concen: 86.729 ug/l

RT: 7.793 min Scan# 1055

Delta R.T. 0.000 min

Lab File: VD079611.D


Acq: 06 Aug 2024 06:55

Instrument :

MSVOA_D

ClientSampleId :

S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion:	97	Resp:	9736
Ion Ratio	100	Lower	Upper
97	100		
99	66.1	51.3	76.9
61	43.0	31.4	47.2

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024

Supervised By :Mahesh
Dadoda
08/08/2024



#33

1,2-Dichloroethane-d4

Concen: 124.296 ug/l




RT: 8.228 min Scan# 1129


Delta R.T. 0.000 min

Lab File: VD079611.D

Acq: 06 Aug 2024 06:55

Tgt Ion:	65	Resp:	79744
Ion Ratio	100	Lower	Upper
65	100		
67	56.3	0.0	111.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD079611.D

Acq: 06 Aug 2024 06:55

Instrument :

MSVOA_D

ClientSampleId :

S-863-K2-SO-12.75-13.25-072024MS

Tgt Ion:114 Resp: 155140





Ion	Ratio	Lower	Upper
114	100		
63	17.8	0.0	34.0
88	12.1	0.0	27.2

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024

Supervised By :Mahesh
Dadoda
08/08/2024



#35

Dibromofluoromethane

Concen: 75.509 ug/l

RT: 7.805 min Scan# 1057




Delta R.T. 0.000 min


Lab File: VD079611.D

Acq: 06 Aug 2024 06:55

Tgt Ion:113 Resp: 79289


Ion	Ratio	Lower	Upper
113	100		
111	102.1	84.0	126.0
192	17.8	15.4	23.0





#36
1,1-Dichloropropene
Concen: 46.389 ug/l
RT: 8.010 min Scan# 1091
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 75 Resp: 66480
Ion Ratio Lower Upper
75 100
110 35.8 19.6 58.7
77 31.8 26.1 39.1

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#37
Ethyl Acetate
Concen: 65.695 ug/l
RT: 7.169 min Scan# 949
Delta R.T. -0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 43 Resp: 38604
Ion Ratio Lower Upper
43 100
61 0.0 13.9 20.9#
70 0.0 9.8 14.6#





#38
Carbon Tetrachloride
Concen: 41.586 ug/l
RT: 7.993 min Scan# 1088
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 117 Resp: 6265
Ion Ratio Lower Upper
117 100
119 101.6 75.4 113.0
121 32.4 25.0 37.4

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#39
Methylcyclohexane
Concen: 22.431 ug/l
RT: 9.275 min Scan# 1307
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 83 Resp: 38358
Ion Ratio Lower Upper
83 100
55 70.1 50.1 75.1
98 50.3 41.9 62.9





#40
Benzene
Concen: 68.598 ug/l
RT: 8.252 min Scan# 1133
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55



Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS




Tgt Ion: 78 Resp: 312269
Ion Ratio Lower Upper
78 100
77 25.4 18.1 27.1

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024







Supervised By :Mahesh
Dadoda
08/08/2024



#41
Methacrylonitrile
Concen: 70.641 ug/l
RT: 7.393 min Scan# 987
Delta R.T. -0.024 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Tgt Ion: 41 Resp: 22833
Ion Ratio Lower Upper
41 100
39 64.8 41.4 62.0#
67 91.9 85.8 128.6
52 34.9 30.4 45.6





#42

1,2-Dichloroethane

Concen: 81.071 ug/l

RT: 8.322 min Scan# 1145

Delta R.T. 0.000 min

Lab File: VD079611.D

Acq: 06 Aug 2024 06:55

Instrument :

MSVOA_D

ClientSampleId :

S-863-K2-SO-12.75-13.25-072024MS





Tgt Ion: 62 Resp: 8794
 Ion Ratio Lower Upper
 62 100
 98 11.4 0.0 25.8

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/06/2024
 Supervised By :Mahesh Dadoda 08/08/2024

Supervised By :Mahesh
 Dadoda
 08/08/2024



#43

Isopropyl Acetate

Concen: 58.783 ug/l




RT: 8.363 min Scan# 1152


Delta R.T. 0.006 min

Lab File: VD079611.D

Acq: 06 Aug 2024 06:55


Tgt Ion: 43 Resp: 64920
 Ion Ratio Lower Upper
 43 100
 61 23.4 29.3 43.9#
 87 15.0 14.9 22.3





#44
Trichloroethene
Concen: 438.743 ug/l
RT: 9.028 min Scan# 1265
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS





Tgt Ion:130 Resp: 49717
Ion Ratio Lower Upper
130 100
95 98.2 0.0 177.8

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#45
1,2-Dichloropropane
Concen: 72.905 ug/l
RT: 9.305 min Scan# 1312
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 63 Resp: 80764
Ion Ratio Lower Upper
63 100
65 32.0 23.8 35.8





#46
Dibromomethane
Concen: 72.924 ug/l
RT: 9.393 min Scan# 1327
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 93 Resp: 4478
Ion Ratio Lower Upper
93 100
95 85.9 65.6 98.4
174 90.7 82.3 123.5

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#47
Bromodichloromethane
Concen: 60.510 ug/l
RT: 9.587 min Scan# 1360
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Tgt Ion: 83 Resp: 91247
Ion Ratio Lower Upper
83 100
85 64.2 50.1 75.1
127 9.3 6.8 10.2





#48

Methyl methacrylate

Concen: 70.408 ug/l

RT: 9.381 min Scan# 1325

Delta R.T. 0.000 min

Lab File: VD079611.D

Acq: 06 Aug 2024 06:55

Instrument :

MSVOA_D

ClientSampleId :

S-863-K2-SO-12.75-13.25-072024MS

Tgt Ion: 41 Resp: 3479

Ion	Ratio	Lower	Upper
41	100		
69	99.6	96.7	145.1
39	51.4	43.8	65.6

Manual Integrations

APPROVED





Reviewed By :Romaben Patel 08/06/2024

Supervised By :Mahesh Dadoda 08/08/2024

Supervised By :Mahesh

Dadoda

08/08/2024



#49

1,4-Dioxane

Concen: 1490.032 ug/l

RT: 9.387 min Scan# 1326




Delta R.T. -0.006 min


Lab File: VD079611.D

Acq: 06 Aug 2024 06:55

Tgt Ion: 88 Resp: 9482


Ion	Ratio	Lower	Upper
88	100		
43	31.3	21.5	32.3
58	58.4	42.2	63.4





#50
Toluene-d8
Concen: 50.596 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55



Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS




Tgt Ion: 98 Resp: 206317
Ion Ratio Lower Upper
98 100
100 64.1 51.6 77.4

Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024





Supervised By :Mahesh
Dadoda
08/08/2024




#51
4-Methyl-2-Pentanone
Concen: 351.360 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55




Tgt Ion: 43 Resp: 198113
Ion Ratio Lower Upper
43 100
58 41.4 36.6 55.0





#52
Toluene
Concen: 53.974 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55




Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS




Tgt Ion: 92 Resp: 151950
Ion Ratio Lower Upper
92 100
91 176.3 141.7 212.5

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024




#53
t-1,3-Dichloropropene
Concen: 44.185 ug/l
RT: 10.552 min Scan# 1524
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55




Tgt Ion: 75 Resp: 61531
Ion Ratio Lower Upper
75 100
77 33.6 25.0 37.6





#54
cis-1,3-Dichloropropene
Concen: 44.583 ug/l
RT: 10.016 min Scan# 1433
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 75 Resp: 74680
Ion Ratio Lower Upper
75 100
77 31.6 24.4 36.6
39 33.7 24.2 36.2

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#55
1,1,2-Trichloroethane
Concen: 70.100 ug/l
RT: 10.728 min Scan# 1554
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 97 Resp: 56711
Ion Ratio Lower Upper
97 100
83 84.8 65.8 98.8
85 52.4 43.8 65.8
99 60.8 55.4 83.0





#56
Ethyl methacrylate
Concen: 58.870 ug/l
RT: 10.599 min Scan# 1532
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 69 Resp: 5643
Ion Ratio Lower Upper
69 100
41 60.0 39.8 59.6
39 28.6 20.2 30.4

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#57
1,3-Dichloropropane
Concen: 67.720 ug/l
RT: 10.875 min Scan# 1579
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Tgt Ion: 76 Resp: 93096
Ion Ratio Lower Upper
76 100
78 32.6 26.2 39.2





#58
2-Chloroethyl Vinyl ether
Concen: 31.622 ug/l
RT: 9.875 min Scan# 1408
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55





Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS

Tgt Ion: 63 Resp: 13979
Ion Ratio Lower Upper
63 100
106 28.3 26.4 39.6

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#59
2-Hexanone
Concen: 263.720 ug/l
RT: 10.922 min Scan# 1587
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 43 Resp: 103705
Ion Ratio Lower Upper
43 100
58 59.4 31.1 93.5





#60
Dibromochloromethane
Concen: 55.090 ug/l
RT: 11.069 min Scan# 1612
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument :
MSVOA_D
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion:129 Resp: 5822
Ion Ratio Lower Upper
129 100
127 76.8 38.0 113.9


Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024



Supervised By :Mahesh
Dadoda
08/08/2024




#61
1,2-Dibromoethane
Concen: 62.773 ug/l
RT: 11.181 min Scan# 1631
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55




Tgt Ion:107 Resp: 47576
Ion Ratio Lower Upper
107 100
109 94.2 72.0 108.0





#62
4-Bromofluorobenzene
Concen: 27.681 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 95 Resp: 32664
Ion Ratio Lower Upper
95 100
174 89.2 0.0 195.4
176 85.0 0.0 192.6

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1699
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion:117 Resp: 110308
Ion Ratio Lower Upper
117 100
82 52.8 39.8 59.6
119 35.4 25.7 38.5





#64
Tetrachloroethene
Concen: 81.278 ug/l
RT: 10.810 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion:164 Resp: 6045
Ion Ratio Lower Upper
164 100
166 120.5 108.6 163.0
129 94.8 74.6 112.0
131 87.0 68.6 103.0

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#65
Chlorobenzene
Concen: 59.995 ug/l
RT: 11.604 min Scan# 1703
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion:112 Resp: 146545
Ion Ratio Lower Upper
112 100
114 30.0 26.0 39.0





#66
1,1,1,2-Tetrachloroethane
Concen: 67.312 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion	Ratio	Lower	Upper
131	100		
133	95.4	49.2	147.6
119	67.6	33.4	100.2

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#67
Ethyl Benzene
Concen: 51.723 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Tgt Ion	Ratio	Lower	Upper
91	100		
106	31.5	25.8	38.6





#68
m/p-Xylenes
Concen: 100.416 ug/l
RT: 11.793 min Scan# 1735
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55





Instrument :
MSVOA_D
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024MS

Tgt Ion:106 Resp: 15930
Ion Ratio Lower Upper
106 100
91 199.0 148.5 222.7

Manual Integrations APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#69
o-Xylene
Concen: 54.918 ug/l
RT: 12.116 min Scan# 1790
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion:106 Resp: 79858
Ion Ratio Lower Upper
106 100
91 211.5 96.7 290.1





#70
Styrene
Concen: 26.239 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion:104 Resp: 66892
Ion Ratio Lower Upper
104 100
78 53.5 33.0 49.6
103 64.2 44.2 66.4

Manual Integrations APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#71
Bromoform
Concen: 65.389 ug/l
RT: 12.293 min Scan# 1820
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion:173 Resp: 29708
Ion Ratio Lower Upper
173 100
175 49.1 23.7 71.1
254 0.0 0.0 0.0





#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion:152 Resp: 27108
Ion Ratio Lower Upper
152 100
115 56.0 26.5 79.5
150 178.8 0.0 341.6

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#73
Isopropylbenzene
Concen: 73.019 ug/l
RT: 12.416 min Scan# 1841
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion:105 Resp: 151905
Ion Ratio Lower Upper
105 100
120 28.0 13.6 40.7





#74
N-amy1 acetate
Concen: 24.965 ug/l
RT: 12.234 min Scan# 1809
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion: 43 Resp: 9954
Ion Ratio Lower Upper
43 100
70 43.1 44.3 66.5
55 27.1 23.2 34.8
61 25.5 24.2 36.4

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#75
1,1,2,2-Tetrachloroethane
Concen: 162.157 ug/l
RT: 12.669 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Tgt Ion: 83 Resp: 57105
Ion Ratio Lower Upper
83 100
131 9.9 5.5 16.5
85 62.1 33.8 101.3





#76
1,2,3-Trichloropropane
Concen: 136.353 ug/l m
RT: 12.722 min Scan# 1893
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55





Instrument :
MSVOA_D
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024MS

Tgt Ion: 75 Resp: 32309
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#77
Bromobenzene
Concen: 100.288 ug/l
RT: 12.698 min Scan# 1889
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Tgt Ion:156 Resp: 45585
Ion Ratio Lower Upper
156 100
77 181.8 77.3 231.9
158 96.1 49.6 149.0









#78
n-propylbenzene
Concen: 69.159 ug/l
RT: 12.757 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS

Tgt Ion: 91 Resp: 156712
Ion Ratio Lower Upper
91 100
120 23.4 12.3 36.8




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024



#79
2-Chlorotoluene
Concen: 83.319 ug/l
RT: 12.845 min Scan# 1914
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 91 Resp: 105126
Ion Ratio Lower Upper
91 100
126 36.0 19.3 57.8





#80
1,3,5-Trimethylbenzene
Concen: 77.841 ug/l
RT: 12.898 min Scan# 1923
Delta R.T. -0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS





Tgt Ion:105 Resp: 11754
Ion Ratio Lower Upper
105 100
120 49.5 25.7 77.1

Manual Integrations APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#81
trans-1,4-Dichloro-2-butene
Concen: 64.144 ug/l
RT: 12.463 min Scan# 1849
Delta R.T. -0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55


Tgt Ion: 75 Resp: 7969
Ion Ratio Lower Upper
75 100
53 77.3 61.4 92.2
89 38.6 35.4 53.0





#82
4-Chlorotoluene
Concen: 77.719 ug/l
RT: 12.945 min Scan# 1931
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55



Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS




Tgt Ion: 91 Resp: 102220
Ion Ratio Lower Upper
91 100
126 36.8 18.9 56.9

Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024





Supervised By :Mahesh
Dadoda
08/08/2024




#83
tert-Butylbenzene
Concen: 62.307 ug/l
RT: 13.163 min Scan# 1968
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55




Tgt Ion:119 Resp: 92543
Ion Ratio Lower Upper
119 100
91 60.2 28.2 84.7
134 24.4 13.4 40.1





#84
1,2,4-Trimethylbenzene
Concen: 66.979 ug/l
RT: 13.210 min Scan# 1976
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55




Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS




Tgt Ion:105 Resp: 115914
Ion Ratio Lower Upper
105 100
120 47.2 23.6 71.0

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024




#85
sec-Butylbenzene
Concen: 52.272 ug/l
RT: 13.340 min Scan# 1998
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55




Tgt Ion:105 Resp: 104988
Ion Ratio Lower Upper
105 100
134 20.2 10.3 30.8





#86
p-Isopropyltoluene
Concen: 45.626 ug/l
RT: 13.457 min Scan# 2018
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion:119 Resp: 8601
Ion Ratio Lower Upper
119 100
134 26.8 13.7 41.1
91 21.2 10.7 32.1

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#87
1,3-Dichlorobenzene
Concen: 65.952 ug/l
RT: 13.457 min Scan# 2018
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Tgt Ion:146 Resp: 62097
Ion Ratio Lower Upper
146 100
111 38.7 18.5 55.5
148 64.2 30.9 92.8





#88
1,4-Dichlorobenzene
Concen: 65.926 ug/l
RT: 13.534 min Scan# 2031
Delta R.T. 0.000 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55





Instrument : MSVOA_D
ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS

Tgt Ion:146	Resp: 61020
Ion Ratio	Lower Upper
146	100
111	37.9 17.5 52.6
148	63.1 31.4 94.2

Manual Integrations
APPROVED




Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh
Dadoda
08/08/2024



#89
n-Butylbenzene
Concen: 31.932 ug/l
RT: 13.787 min Scan# 2074
Delta R.T. 0.006 min
Lab File: VD079611.D
Acq: 06 Aug 2024 06:55

Tgt Ion: 91	Resp: 53446
Ion Ratio	Lower Upper
91	100
92	49.5 25.7 77.0
134	26.1 13.6 40.8





#90

Hexachloroethane

Concen: 53.337 ug/l

RT: 14.051 min Scan# 2119

Delta R.T. 0.006 min

Lab File: VD079611.D

Acq: 06 Aug 2024 06:55

Instrument :

MSVOA_D

ClientSampleId :

S-863-K2-SO-12.75-13.25-072024MS





Tgt Ion:117 Resp: 1873
 Ion Ratio Lower Upper
 117 100
 201 71.2 36.3 108.9

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/06/2024
 Supervised By :Mahesh Dadoda 08/08/2024

Supervised By :Mahesh
 Dadoda
 08/08/2024



#91

1,2-Dichlorobenzene

Concen: 73.112 ug/l




RT: 13.828 min Scan# 2081


Delta R.T. 0.000 min

Lab File: VD079611.D

Acq: 06 Aug 2024 06:55


Tgt Ion:146 Resp: 58832
 Ion Ratio Lower Upper
 146 100
 111 39.8 19.1 57.1
 148 62.2 31.1 93.5





#92
 1,2-Dibromo-3-Chloropropane
 Concen: 127.193 ug/l
 RT: 14.440 min Scan# 2185
 Delta R.T. -0.006 min
 Lab File: VD079611.D
 Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
 ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS






Tgt Ion:	75	Resp:	5960
Ion Ratio	Lower	Upper	
75	100		
155	83.2	47.7	143.1
157	115.5	63.4	190.2

Manual Integrations
 APPROVED




Reviewed By :Romaben Patel 08/06/2024
 Supervised By :Mahesh Dadoda 08/08/2024


Supervised By :Mahesh Dadoda
 08/08/2024



#93
 1,2,4-Trichlorobenzene
 Concen: 28.377 ug/l
 RT: 15.098 min Scan# 2297
 Delta R.T. 0.006 min
 Lab File: VD079611.D
 Acq: 06 Aug 2024 06:55

Tgt Ion:	180	Resp:	14387
Ion Ratio	Lower	Upper	
180	100		
182	90.8	46.9	140.7
145	28.9	14.0	41.9





#94

Hexachlorobutadiene

Concen: 16.939 ug/l

RT: 15.198 min Scan# 21

Delta R.T. 0.000 min

Lab File: VD079611.D

Acq: 06 Aug 2024 06:55

Instrument :

MSVOA_D

ClientSampleId :

S-863-K2-SO-12.75-13.25-072024MS





Tgt Ion:	225	Resp:	463
Ion Ratio	Lower	Upper	
225	100		
223	64.2	33.3	99.9
227	60.2	33.6	100.6

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/06/2024
Supervised By :Mahesh Dadoda 08/08/2024

Supervised By :Mahesh
Dadoda
08/08/2024



#95

Naphthalene

Concen: 50.126 ug/l




RT: 15.322 min Scan# 2335


Delta R.T. -0.006 min

Lab File: VD079611.D

Acq: 06 Aug 2024 06:55


Tgt Ion:	128	Resp:	43986
Ion Ratio	Lower	Upper	
128	100		
127	14.4	10.2	15.4
129	12.0	8.5	12.7





#96
 1,2,3-Trichlorobenzene
 Concen: 29.506 ug/l
 RT: 15.516 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: VD079611.D
 Acq: 06 Aug 2024 06:55

Instrument : MSVOA_D
 ClientSampleId : S-863-K2-SO-12.75-13.25-072024MS



Tgt Ion	Ratio	Lower	Upper
180	100		
182	92.5	47.8	143.3
145	30.4	16.1	48.2

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 08/06/2024
 Supervised By :Mahesh Dadoda 08/08/2024

Supervised By :Mahesh
 Dadoda
 08/08/2024

