



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080524\
Data File : VD079593.D
Acq On : 05 Aug 2024 21:23
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Title : SW846 8260
Last Update : Tue Jul 30 03:50:47 2024



AutoFind: Scans 1866, 1867, 1868; Background Corrected with Scan 1857

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	26139	PASS
75	95	30	60	50.3	69291	PASS
95	95	100	100	100.0	137792	PASS
96	95	5	9	6.2	8543	PASS
173	174	0.00	2	1.5	1827	PASS
174	95	50	100	89.1	122728	PASS
175	174	5	9	7.6	9314	PASS
176	174	95	101	97.1	119221	PASS
177	176	5	9	6.9	8182	PASS