

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080724\
 Data File : VD079646.D
 Acq On : 07 Aug 2024 17:36
 Operator : RP/MD
 Sample : P3393-08
 Misc : 8.40G/5.0ml/MSVOA_D/SOILA
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-851-GI-SO-10-10.5-072624

Quant Time: Aug 08 01:53:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	163207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	309742	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	279385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	118041	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	95825	57.795	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.580%
35) Dibromofluoromethane	7.810	113	107212	51.139	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.280%
50) Toluene-d8	10.269	98	376983	46.305	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.600%
62) 4-Bromofluorobenzene	12.575	95	97573	41.415	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.840%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.281	62	26889	9.598	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.964	101	11087	6.292	ug/l	94
16) Acetone	4.034	43	1691	5.248	ug/l #	80
21) trans-1,2-Dichloroethene	5.311	96	5662	2.858	ug/l	97
27) cis-1,2-Dichloroethene	7.075	96	659693	291.339	ug/l	99
44) Trichloroethene	9.028	130	1221529	539.917	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080724\
Data File : VD079646.D
Acq On : 07 Aug 2024 17:36
Operator : RP/MD
Sample : P3393-08
Misc : 8.40G/5.0ml/MSVOA_D/SOILA
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-851-GI-SO-10-10.5-072624

Quant Time: Aug 08 01:53:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

