

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080824\
 Data File : VD079671.D
 Acq On : 08 Aug 2024 17:08
 Operator : RP/MD
 Sample : P3393-01RE
 Misc : 6.54G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-848-GI-SO-2.5-3-072624RE

Quant Time: Aug 09 01:09:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	185681	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	371652	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	290517	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	90136	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	94726	50.217	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.440%
35) Dibromofluoromethane	7.804	113	107858	42.877	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.760%
50) Toluene-d8	10.269	98	385486	39.462	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	78.920%
62) 4-Bromofluorobenzene	12.575	95	80236	28.383	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	56.760%
Target Compounds						Qvalue
44) Trichloroethene	9.028	130	54811	20.191	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080824\
Data File : VD079671.D
Acq On : 08 Aug 2024 17:08
Operator : RP/MD
Sample : P3393-01RE
Misc : 6.54G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-848-GI-SO-2.5-3-072624RE

Quant Time: Aug 09 01:09:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

