

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080824\
 Data File : VD079672.D
 Acq On : 08 Aug 2024 17:35
 Operator : RP/MD
 Sample : P3393-07RE
 Misc : 8.77G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-850-GI-SO-2.5-3-072624RE

Quant Time: Aug 09 01:10:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
 Supervised By :Semsettin Yesilyurt 08/09/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	185356	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	369949	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	289194	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	81607	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	98504	52.311	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 104.620%			
35) Dibromofluoromethane	7.804	113	108423	43.300	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 86.600%			
50) Toluene-d8	10.269	98	375895	38.657	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 77.320%			
62) 4-Bromofluorobenzene	12.575	95	77182	27.429	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 54.860%			
Target Compounds						
16) Acetone	4.010	43	3342m	9.132	ug/l	
17) Carbon Disulfide	4.263	76	7854	1.188	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080824\
Data File : VD079672.D
Acq On : 08 Aug 2024 17:35
Operator : RP/MD
Sample : P3393-07RE
Misc : 8.77G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-850-GI-SO-2.5-3-072624RE

Quant Time: Aug 09 01:10:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024

