

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080824\
 Data File : VD079667.D
 Acq On : 08 Aug 2024 15:17
 Operator : RP/MD
 Sample : VD0808SBS01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0808SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
 Supervised By :Semsettin Yesilyurt 08/09/2024

Quant Time: Aug 09 01:07:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	174894	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	330058	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	299617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	148188	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	103047	57.997	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	116.000%		
35) Dibromofluoromethane	7.799	113	105537	47.242	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	94.480%		
50) Toluene-d8	10.269	98	416540	48.014	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	96.020%		
62) 4-Bromofluorobenzene	12.569	95	121040	48.214	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	96.420%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.934	85	29183	21.296	ug/l	99
3) Chloromethane	2.146	50	58249	23.070	ug/l	100
4) Vinyl Chloride	2.281	62	70588	23.513	ug/l	100
5) Bromomethane	2.675	94	52739	26.599	ug/l	97
6) Chloroethane	2.828	64	49808	25.819	ug/l	100
7) Trichlorofluoromethane	3.169	101	70021	23.453	ug/l	99
8) Diethyl Ether	3.593	74	23868	26.465	ug/l	79
9) 1,1,2-Trichlorotrifluo...	3.958	101	41455	21.954	ug/l	97
10) Methyl Iodide	4.158	142	39903	18.906	ug/l	96
11) Tert butyl alcohol	5.058	59	15994	202.394	ug/l #	82
12) 1,1-Dichloroethene	3.934	96	42604	22.545	ug/l	87
13) Acrolein	3.805	56	16030	123.444	ug/l	99
14) Allyl chloride	4.564	41	66503	26.651	ug/l #	83
15) Acrylonitrile	5.246	53	61586	146.135	ug/l	97
16) Acetone	4.022	43	40183	116.372	ug/l	96
17) Carbon Disulfide	4.264	76	112213	17.990	ug/l	98
18) Methyl Acetate	4.569	43	32167	30.004	ug/l #	87
19) Methyl tert-butyl Ether	5.316	73	104301	27.040	ug/l	97
20) Methylene Chloride	4.799	84	73989	20.835	ug/l #	81
21) trans-1,2-Dichloroethene	5.311	96	48463	22.829	ug/l #	82
22) Diisopropyl ether	6.210	45	156718	28.227	ug/l #	89
23) Vinyl Acetate	6.158	43	586840	140.607	ug/l #	90
24) 1,1-Dichloroethane	6.110	63	98558	26.489	ug/l	97
25) 2-Butanone	7.081	43	69123	141.201	ug/l	92
26) 2,2-Dichloropropane	7.069	77	69536	23.105	ug/l	100
27) cis-1,2-Dichloroethene	7.081	96	60572	24.963	ug/l	85
28) Bromochloromethane	7.428	49	41357	25.348	ug/l #	65
29) Tetrahydrofuran	7.446	42	45175	148.906	ug/l #	80
30) Chloroform	7.599	83	99120	26.307	ug/l	99
31) Cyclohexane	7.881	56	63406	21.487	ug/l	84
32) 1,1,1-Trichloroethane	7.793	97	75776	24.372	ug/l	96
36) 1,1-Dichloropropene	8.005	75	62692	20.560	ug/l	95
37) Ethyl Acetate	7.175	43	33569	26.852	ug/l #	91
38) Carbon Tetrachloride	7.993	117	65659	20.484	ug/l	99
39) Methylcyclohexane	9.275	83	67803	18.637	ug/l	91
40) Benzene	8.252	78	211631	21.852	ug/l	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080824\
 Data File : VD079667.D
 Acq On : 08 Aug 2024 15:17
 Operator : RP/MD
 Sample : VD0808SBS01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0808SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
 Supervised By :Semsettin Yesilyurt 08/09/2024

Quant Time: Aug 09 01:07:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.399	41	16350	23.776	ug/l #	84
42) 1,2-Dichloroethane	8.322	62	53780	23.304	ug/l	97
43) Isopropyl Acetate	8.363	43	58506	24.901	ug/l	94
44) Trichloroethene	9.034	130	49833	20.670	ug/l	89
45) 1,2-Dichloropropane	9.304	63	54752	23.231	ug/l	98
46) Dibromomethane	9.393	93	28603	21.892	ug/l	91
47) Bromodichloromethane	9.587	83	73975	23.058	ug/l	96
48) Methyl methacrylate	9.381	41	26607	25.304	ug/l #	79
49) 1,4-Dioxane	9.381	88	6486	479.078	ug/l #	85
51) 4-Methyl-2-Pentanone	10.157	43	160062	133.432	ug/l	93
52) Toluene	10.334	92	131195	21.904	ug/l	100
53) t-1,3-Dichloropropene	10.551	75	67803	22.886	ug/l	100
54) cis-1,3-Dichloropropene	10.016	75	80337	22.543	ug/l #	91
55) 1,1,2-Trichloroethane	10.728	97	40864	23.742	ug/l	93
56) Ethyl methacrylate	10.593	69	48597	23.830	ug/l #	87
57) 1,3-Dichloropropane	10.875	76	70116	23.974	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.869	63	97498	103.668	ug/l	90
59) 2-Hexanone	10.916	43	108005	129.098	ug/l	90
60) Dibromochloromethane	11.075	129	49009	21.797	ug/l	98
61) 1,2-Dibromoethane	11.181	107	34971	21.688	ug/l	95
64) Tetrachloroethene	10.810	164	38650	19.131	ug/l	95
65) Chlorobenzene	11.604	112	146317	22.053	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.675	131	49056	22.473	ug/l	96
67) Ethyl Benzene	11.681	91	233812	21.527	ug/l	99
68) m/p-Xylenes	11.787	106	186496	43.280	ug/l	89
69) o-Xylene	12.122	106	88832	22.491	ug/l	93
70) Styrene	12.134	104	155000	22.384	ug/l	96
71) Bromoform	12.293	173	26830	21.742	ug/l #	99
73) Isopropylbenzene	12.416	105	223350	22.292	ug/l	99
74) N-amyl acetate	12.234	43	59530	27.312	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	12.669	83	47935	24.900	ug/l	93
76) 1,2,3-Trichloropropane	12.722	75	34936m	26.971	ug/l	
77) Bromobenzene	12.698	156	52266	21.034	ug/l	82
78) n-propylbenzene	12.757	91	282482	22.805	ug/l	97
79) 2-Chlorotoluene	12.845	91	157243	22.798	ug/l	95
80) 1,3,5-Trimethylbenzene	12.898	105	187736	22.743	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.469	75	16326	24.039	ug/l	95
82) 4-Chlorotoluene	12.945	91	164547	22.885	ug/l	96
83) tert-Butylbenzene	13.163	119	157135	21.854	ug/l	95
84) 1,2,4-Trimethylbenzene	13.210	105	187142	22.197	ug/l	99
85) sec-Butylbenzene	13.339	105	245246	22.337	ug/l	99
86) p-Isopropyltoluene	13.457	119	199852	21.626	ug/l	98
87) 1,3-Dichlorobenzene	13.457	146	108290	21.039	ug/l	96
88) 1,4-Dichlorobenzene	13.534	146	109512	21.641	ug/l	97
89) n-Butylbenzene	13.787	91	183759	21.487	ug/l	97
90) Hexachloroethane	14.051	117	41871	21.803	ug/l	93
91) 1,2-Dichlorobenzene	13.828	146	95882	21.797	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	14.445	75	6500	25.376	ug/l	81
93) 1,2,4-Trichlorobenzene	15.098	180	53708	19.378	ug/l	99
94) Hexachlorobutadiene	15.198	225	27289	18.243	ug/l	98
95) Naphthalene	15.328	128	109745	22.878	ug/l	98
96) 1,2,3-Trichlorobenzene	15.516	180	49048	20.118	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080824\
Data File : VD079667.D
Acq On : 08 Aug 2024 15:17
Operator : RP/MD
Sample : VD0808SBS01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024

Quant Time: Aug 09 01:07:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080824\
Data File : VD079667.D
Acq On : 08 Aug 2024 15:17
Operator : RP/MD
Sample : VD0808SBS01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01

Manual Integrations
APPROVED

Reviewed By : Mahesh Dadoda 08/09/2024
Supervised By : Semsettin Yesilyurt 08/09/2024


Quant Time: Aug 09 01:07:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 10
Delta R.T. 0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion:168 Resp: 174894
Ion Ratio Lower Upper
168 100
99 57.7 41.1 61.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#2
Dichlorodifluoromethane
Concen: 21.296 ug/l
RT: 1.934 min Scan# 59
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion: 85 Resp: 29183
Ion Ratio Lower Upper
85 100
87 32.4 16.0 48.0





#3
Chloromethane
Concen: 23.070 ug/l
RT: 2.146 min Scan# 91
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01




Tgt Ion: 50 Resp: 58249
Ion Ratio Lower Upper
50 100
52 35.1 28.2 42.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024




#4
Vinyl Chloride
Concen: 23.513 ug/l
RT: 2.281 min Scan# 118
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Tgt Ion: 62 Resp: 70588
Ion Ratio Lower Upper
62 100
64 33.1 26.3 39.5





#5
Bromomethane
Concen: 26.599 ug/l
RT: 2.675 min Scan# 11
Delta R.T. -0.012 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion: 94 Resp: 52739
Ion Ratio Lower Upper
94 100
96 93.6 76.9 115.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#6
Chloroethane
Concen: 25.819 ug/l
RT: 2.828 min Scan# 211
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion: 64 Resp: 49808
Ion Ratio Lower Upper
64 100
66 30.7 24.5 36.7





#7
 Trichlorofluoromethane
 Concen: 23.453 ug/l
 RT: 3.169 min Scan# 218
 Delta R.T. -0.000 min
 Lab File: VD079667.D
 Acq: 08 Aug 2024 15:17




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0808SBS01




Tgt Ion:101 Resp: 7002
 Ion Ratio Lower Upper
 101 100
 103 67.5 53.2 79.8

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
 Supervised By :Semsettin Yesilyurt 08/09/2024




#8
 Diethyl Ether
 Concen: 26.465 ug/l
 RT: 3.593 min Scan# 341
 Delta R.T. -0.000 min
 Lab File: VD079667.D
 Acq: 08 Aug 2024 15:17



Tgt Ion: 74 Resp: 23868
 Ion Ratio Lower Upper
 74 100
 45 93.6 37.9 113.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.954 ug/l
RT: 3.958 min Scan# 403
Delta R.T. -0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01

Tgt Ion:101 Resp: 4145
Ion Ratio Lower Upper
101 100
85 44.9 33.5 50.3
151 74.1 60.3 90.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#10
Methyl Iodide
Concen: 18.906 ug/l
RT: 4.158 min Scan# 437
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Tgt Ion:142 Resp: 39903
Ion Ratio Lower Upper
142 100
127 45.3 34.2 51.4
141 16.8 11.7 17.5





#11

Tert butyl alcohol

Concen: 202.394 ug/l

RT: 5.058 min Scan# 513

Delta R.T. 0.012 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :

VD0808SBS01

Tgt Ion: 59 Resp: 15994

Ion Ratio Lower Upper

59 100





57 10.4 3.6 5.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024

Supervised By :Semsettin Yesilyurt 08/09/2024



#12

1,1-Dichloroethene

Concen: 22.545 ug/l

RT: 3.934 min Scan# 399

Delta R.T. 0.006 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17




Tgt Ion: 96 Resp: 42604


Ion Ratio Lower Upper

96 100

61 152.7 107.3 160.9


98 59.1 51.7 77.5





#13
Acrolein
Concen: 123.444 ug/l
RT: 3.805 min Scan# 316
Delta R.T. 0.012 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion: 56 Resp: 16030
Ion Ratio Lower Upper
56 100
55 68.4 55.3 82.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#14
Allyl chloride
Concen: 26.651 ug/l
RT: 4.564 min Scan# 506
Delta R.T. 0.012 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion: 41 Resp: 66503
Ion Ratio Lower Upper
41 100
39 60.4 58.0 87.0
76 42.9 45.9 68.9#





#15
Acrylonitrile
Concen: 146.135 ug/l
RT: 5.246 min Scan# 61580
Delta R.T. -0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01




Tgt Ion: 53 Resp: 61580
Ion Ratio Lower Upper
53 100
52 83.1 67.7 101.5
51 38.4 28.6 42.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024




#16
Acetone
Concen: 116.372 ug/l
RT: 4.022 min Scan# 414
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17



Tgt Ion: 43 Resp: 40183
Ion Ratio Lower Upper
43 100
58 35.8 30.4 45.6









#17
Carbon Disulfide
Concen: 17.990 ug/l
RT: 4.264 min Scan# 411
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01

Tgt Ion: 76 Resp: 11221
Ion Ratio Lower Upper
76 100
78 9.3 6.9 10.3




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#18
Methyl Acetate
Concen: 30.004 ug/l
RT: 4.569 min Scan# 507
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion: 43 Resp: 32167
Ion Ratio Lower Upper
43 100
74 27.4 28.1 42.1#





#19
Methyl tert-butyl Ether
Concen: 27.040 ug/l
RT: 5.316 min Scan# 61
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01




Tgt Ion: 73 Resp: 10430
Ion Ratio Lower Upper
73 100
57 24.6 18.3 27.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#20
Methylene Chloride
Concen: 20.835 ug/l
RT: 4.799 min Scan# 546
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Tgt Ion: 84 Resp: 73989
Ion Ratio Lower Upper
84 100
49 122.2 75.6 113.4#
51 35.6 24.2 36.2
86 67.2 49.2 73.8





#21
trans-1,2-Dichloroethene
Concen: 22.829 ug/l
RT: 5.311 min Scan# 611
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01




Tgt Ion: 96 Resp: 4846
Ion Ratio Lower Upper
96 100
61 135.5 84.5 126.7
98 62.5 50.5 75.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024




#22
Diisopropyl ether
Concen: 28.227 ug/l
RT: 6.210 min Scan# 786
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Tgt Ion: 45 Resp: 156718
Ion Ratio Lower Upper
45 100
43 57.3 41.7 62.5
87 26.9 30.2 45.2#
59 11.1 11.0 16.6





#23
 Vinyl Acetate
 Concen: 140.607 ug/l
 RT: 6.158 min Scan# 711
 Delta R.T. 0.006 min
 Lab File: VD079667.D
 Acq: 08 Aug 2024 15:17




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0808SBS01




Tgt Ion: 43 Resp: 586840
 Ion Ratio Lower Upper
 43 100
 86 11.0 12.3 18.5

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
 Supervised By :Semsettin Yesilyurt 08/09/2024




#24
 1,1-Dichloroethane
 Concen: 26.489 ug/l
 RT: 6.110 min Scan# 769
 Delta R.T. 0.006 min
 Lab File: VD079667.D
 Acq: 08 Aug 2024 15:17



Tgt Ion: 63 Resp: 98558
 Ion Ratio Lower Upper
 63 100
 98 7.8 4.7 14.0
 100 5.1 2.9 8.6





#25

2-Butanone

Concen: 141.201 ug/l

RT: 7.081 min Scan# 912

Delta R.T. -0.012 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :

VD0808SBS01

Tgt Ion: 43 Resp: 6912

Ion Ratio Lower Upper

43 100





72 28.5 26.4 39.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024

Supervised By :Semsettin Yesilyurt 08/09/2024



#26

2,2-Dichloropropane

Concen: 23.105 ug/l

RT: 7.069 min Scan# 932

Delta R.T. -0.012 min

Lab File: VD079667.D




Acq: 08 Aug 2024 15:17


Tgt Ion: 77 Resp: 69536

Ion Ratio Lower Upper

77 100

97 24.9 12.4 37.2





#27

cis-1,2-Dichloroethene

Concen: 24.963 ug/l

RT: 7.081 min Scan# 91

Delta R.T. 0.006 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :

VD0808SBS01

Tgt Ion: 96 Resp: 60572

Ion Ratio Lower Upper

96 100

61 138.7 0.0 229.0





98 64.6 0.0 127.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024

Supervised By :Semsettin Yesilyurt 08/09/2024



#28

Bromochloromethane

Concen: 25.348 ug/l

RT: 7.428 min Scan# 993

Delta R.T. 0.012 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17




Tgt Ion: 49 Resp: 41357

Ion Ratio Lower Upper

49 100

129 0.4 0.0 7.0

130 71.4 87.7 131.5#





#29

Tetrahydrofuran

Concen: 148.906 ug/l

RT: 7.446 min Scan# 907

Delta R.T. -0.000 min

Lab File: VD079667.D


Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :

VD0808SBS01






Tgt Ion:	42	Resp:	4517
Ion Ratio	Lower	Upper	
42	100		
72	45.7	49.4	74.0
71	41.8	45.0	67.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#30

Chloroform

Concen: 26.307 ug/l




RT: 7.599 min Scan# 1022


Delta R.T. 0.006 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17


Tgt Ion:	83	Resp:	99120
Ion Ratio	Lower	Upper	
83	100		
85	65.1	51.8	77.6





#31
Cyclohexane
Concen: 21.487 ug/l
RT: 7.881 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion: 56 Resp: 63400
Ion Ratio Lower Upper
56 100
69 37.7 28.0 42.0
84 88.3 87.4 131.0




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#32
1,1,1-Trichloroethane
Concen: 24.372 ug/l
RT: 7.793 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion: 97 Resp: 75776
Ion Ratio Lower Upper
97 100
99 66.6 51.3 76.9
61 43.2 31.4 47.2





#33
1,2-Dichloroethane-d4
Concen: 57.997 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01




Tgt Ion: 65 Resp: 10304
Ion Ratio Lower Upper
65 100
67 55.6 0.0 111.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.775 min Scan# 1222
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Tgt Ion:114 Resp: 330058
Ion Ratio Lower Upper
114 100
63 19.9 0.0 34.0
88 14.2 0.0 27.2





#35
Dibromofluoromethane
Concen: 47.242 ug/l
RT: 7.799 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion:113 Resp: 10553
Ion Ratio Lower Upper
113 100
111 105.0 84.0 126.0
192 16.7 15.4 23.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#36
1,1-Dichloropropene
Concen: 20.560 ug/l
RT: 8.005 min Scan# 1091
Delta R.T. -0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion: 75 Resp: 62692
Ion Ratio Lower Upper
75 100
110 35.2 19.6 58.7
77 30.4 26.1 39.1





#37
Ethyl Acetate
Concen: 26.852 ug/l
RT: 7.175 min Scan# 91
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion: 43 Resp: 33569
Ion Ratio Lower Upper
43 100
61 12.7 13.9 20.9
70 9.3 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#38
Carbon Tetrachloride
Concen: 20.484 ug/l
RT: 7.993 min Scan# 1089
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Tgt Ion:117 Resp: 65659
Ion Ratio Lower Upper
117 100
119 93.8 75.4 113.0
121 32.0 25.0 37.4





#39

Methylcyclohexane

Concen: 18.637 ug/l

RT: 9.275 min Scan# 1133

Delta R.T. -0.000 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :





VD0808SBS01

Tgt Ion: 83 Resp: 6780
 Ion Ratio Lower Upper
 83 100
 55 72.0 50.1 75.1
 98 49.1 41.9 62.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
 Supervised By :Semsettin Yesilyurt 08/09/2024



#40

Benzene

Concen: 21.852 ug/l




RT: 8.252 min Scan# 1133


Delta R.T. -0.000 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17


Tgt Ion: 78 Resp: 211631
 Ion Ratio Lower Upper
 78 100
 77 25.8 18.1 27.1





#41
Methacrylonitrile
Concen: 23.776 ug/l
RT: 7.399 min Scan# 91
Delta R.T. -0.018 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion: 41 Resp: 16350
Ion Ratio Lower Upper
41 100
39 59.8 41.4 62.0
67 82.4 85.8 128.6
52 37.7 30.4 45.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#42
1,2-Dichloroethane
Concen: 23.304 ug/l
RT: 8.322 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion: 62 Resp: 53780
Ion Ratio Lower Upper
62 100
98 11.5 0.0 25.8





#43
Isopropyl Acetate
Concen: 24.901 ug/l
RT: 8.363 min Scan# 1152
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01




Tgt Ion: 43 Resp: 58500
Ion Ratio Lower Upper
43 100
61 33.1 29.3 43.9
87 15.3 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024




#44
Trichloroethene
Concen: 20.670 ug/l
RT: 9.034 min Scan# 1266
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17



Tgt Ion:130 Resp: 49833
Ion Ratio Lower Upper
130 100
95 99.3 0.0 177.8





#45

1,2-Dichloropropane

Concen: 23.231 ug/l

RT: 9.304 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :

VD0808SBS01

Tgt Ion: 63 Resp: 5475

Ion Ratio Lower Upper





63 100

65 28.5 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#46

Dibromomethane

Concen: 21.892 ug/l

RT: 9.393 min Scan# 1327

Delta R.T. -0.000 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17




Tgt Ion: 93 Resp: 28603


Ion Ratio Lower Upper

93 100

95 86.1 65.6 98.4

174 90.6 82.3 123.5









#47
Bromodichloromethane
Concen: 23.058 ug/l
RT: 9.587 min Scan# 1360
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01

Tgt Ion: 83 Resp: 7397
Ion Ratio Lower Upper
83 100
85 66.0 50.1 75.1
127 7.7 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#48
Methyl methacrylate
Concen: 25.304 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Tgt Ion: 41 Resp: 26607
Ion Ratio Lower Upper
41 100
69 91.6 96.7 145.1#
39 48.1 43.8 65.6





#49

1,4-Dioxane

Concen: 479.078 ug/l

RT: 9.381 min Scan# 11

Delta R.T. -0.012 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :

VD0808SBS01

Tgt Ion: 88 Resp: 6480

Ion Ratio Lower Upper

88 100





43 30.3 21.5 32.3

58 66.4 42.2 63.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#50

Toluene-d8

Concen: 48.014 ug/l

RT: 10.269 min Scan# 1476

Delta R.T. 0.006 min

Lab File: VD079667.D




Acq: 08 Aug 2024 15:17


Tgt Ion: 98 Resp: 416540

Ion Ratio Lower Upper

98 100

100 64.1 51.6 77.4









#51
4-Methyl-2-Pentanone
Concen: 133.432 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01

Tgt Ion: 43 Resp: 160062
Ion Ratio Lower Upper
43 100
58 41.0 36.6 55.0




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#52
Toluene
Concen: 21.904 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion: 92 Resp: 131195
Ion Ratio Lower Upper
92 100
91 177.2 141.7 212.5





#53
t-1,3-Dichloropropene
Concen: 22.886 ug/l
RT: 10.551 min Scan# 1423
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01




Tgt Ion: 75 Resp: 6780
Ion Ratio Lower Upper
75 100
77 31.4 25.0 37.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024




#54
cis-1,3-Dichloropropene
Concen: 22.543 ug/l
RT: 10.016 min Scan# 1433
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Tgt Ion: 75 Resp: 80337
Ion Ratio Lower Upper
75 100
77 28.2 24.4 36.6
39 37.5 24.2 36.2#





#55
1,1,2-Trichloroethane
Concen: 23.742 ug/l
RT: 10.728 min Scan# 1554
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01




Tgt Ion:	97	Resp:	40864
Ion	Ratio	Lower	Upper
97	100		
83	82.9	65.8	98.8
85	51.7	43.8	65.8
99	55.7	55.4	83.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024




#56
Ethyl methacrylate
Concen: 23.830 ug/l
RT: 10.593 min Scan# 1531
Delta R.T. -0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Tgt Ion:	69	Resp:	48597
Ion	Ratio	Lower	Upper
69	100		
41	60.6	39.8	59.6#
39	29.2	20.2	30.4





#57
1,3-Dichloropropane
Concen: 23.974 ug/l
RT: 10.875 min Scan# 110
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion: 76 Resp: 70110
Ion Ratio Lower Upper
76 100
78 31.6 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#58
2-Chloroethyl Vinyl ether
Concen: 103.668 ug/l
RT: 9.869 min Scan# 1408
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion: 63 Resp: 97498
Ion Ratio Lower Upper
63 100
106 27.4 26.4 39.6





#59
2-Hexanone
Concen: 129.098 ug/l
RT: 10.916 min Scan# 1586
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion: 43 Resp: 108009
Ion Ratio Lower Upper
43 100
58 54.9 31.1 93.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#60
Dibromochloromethane
Concen: 21.797 ug/l
RT: 11.075 min Scan# 1613
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Tgt Ion:129 Resp: 49009
Ion Ratio Lower Upper
129 100
127 77.9 38.0 113.9









#61
1,2-Dibromoethane
Concen: 21.688 ug/l
RT: 11.181 min Scan# 1107
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01

Tgt Ion:107 Resp: 3497
Ion Ratio Lower Upper
107 100
109 94.8 72.0 108.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#62
4-Bromofluorobenzene
Concen: 48.214 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion: 95 Resp: 121040
Ion Ratio Lower Upper
95 100
174 83.9 0.0 195.4
176 78.5 0.0 192.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion:117 Resp: 29961
Ion Ratio Lower Upper
117 100
82 57.3 39.8 59.6
119 32.0 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#64
Tetrachloroethene
Concen: 19.131 ug/l
RT: 10.810 min Scan# 1568
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion:164 Resp: 38650
Ion Ratio Lower Upper
164 100
166 128.1 108.6 163.0
129 98.3 74.6 112.0
131 84.3 68.6 103.0





#65
Chlorobenzene
Concen: 22.053 ug/l
RT: 11.604 min Scan# 1715
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion:112 Resp: 14631
Ion Ratio Lower Upper
112 100
114 32.6 26.0 39.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 22.473 ug/l
RT: 11.675 min Scan# 1715
Delta R.T. -0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Tgt Ion:131 Resp: 49056
Ion Ratio Lower Upper
131 100
133 93.7 49.2 147.6
119 65.4 33.4 100.2





#67

Ethyl Benzene

Concen: 21.527 ug/l

RT: 11.681 min Scan# 1716

Delta R.T. -0.000 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :

VD0808SBS01

Tgt Ion: 91 Resp: 23381

Ion Ratio Lower Upper

91 100





106 33.0 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024

Supervised By :Semsettin Yesilyurt 08/09/2024



#68

m/p-Xylenes

Concen: 43.280 ug/l

RT: 11.787 min Scan# 1734

Delta R.T. -0.006 min

Lab File: VD079667.D




Acq: 08 Aug 2024 15:17


Tgt Ion:106 Resp: 186496

Ion Ratio Lower Upper

106 100

91 201.1 148.5 222.7









#69
o-Xylene
Concen: 22.491 ug/l
RT: 12.122 min Scan# 1791
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01

Tgt Ion:106 Resp: 88832
Ion Ratio Lower Upper
106 100
91 204.0 96.7 290.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#70
Styrene
Concen: 22.384 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion:104 Resp: 155000
Ion Ratio Lower Upper
104 100
78 45.8 33.0 49.6
103 56.0 44.2 66.4





#71
Bromoform
Concen: 21.742 ug/l
RT: 12.293 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion:173 Resp: 26830
Ion Ratio Lower Upper
173 100
175 47.9 23.7 71.1
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Tgt Ion:152 Resp: 148188
Ion Ratio Lower Upper
152 100
115 56.7 26.5 79.5
150 160.1 0.0 341.6





#73

Isopropylbenzene

Concen: 22.292 ug/l

RT: 12.416 min Scan# 1810

Delta R.T. 0.000 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :

VD0808SBS01

Tgt Ion:105 Resp: 223350

Ion Ratio Lower Upper

105 100





120 26.6 13.6 40.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024

Supervised By :Semsettin Yesilyurt 08/09/2024



#74

N-ethyl acetate

Concen: 27.312 ug/l

RT: 12.234 min Scan# 1810

Delta R.T. -0.000 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Tgt Ion: 43 Resp: 59530




Ion Ratio Lower Upper


43 100

70 43.3 44.3 66.5#

55 25.0 23.2 34.8


61 25.7 24.2 36.4





#75
1,1,2,2-Tetrachloroethane
Concen: 24.900 ug/l
RT: 12.669 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion: 83 Resp: 47938
Ion Ratio Lower Upper
83 100
131 9.4 5.5 16.5
85 61.6 33.8 101.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#76
1,2,3-Trichloropropane
Concen: 26.971 ug/l m
RT: 12.722 min Scan# 1893
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Tgt Ion: 75 Resp: 34936
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 21.034 ug/l

RT: 12.698 min Scan# 1889

Delta R.T. -0.000 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :





VD0808SBS01

Tgt Ion:	156	Resp:	52260
Ion Ratio	Lower	Upper	
156	100		
77	189.9	77.3	231.9
158	97.0	49.6	149.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#78

n-propylbenzene

Concen: 22.805 ug/l




RT: 12.757 min Scan# 1899


Delta R.T. -0.000 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Tgt Ion:	91	Resp:	282482
Ion Ratio	Lower	Upper	
91	100		
120	22.8	12.3	36.8









#79
2-Chlorotoluene
Concen: 22.798 ug/l
RT: 12.845 min Scan# 1914
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01

Tgt Ion: 91 Resp: 15724
Ion Ratio Lower Upper
91 100
126 35.4 19.3 57.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#80
1,3,5-Trimethylbenzene
Concen: 22.743 ug/l
RT: 12.898 min Scan# 1923
Delta R.T. -0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion:105 Resp: 187736
Ion Ratio Lower Upper
105 100
120 50.4 25.7 77.1





#81
trans-1,4-Dichloro-2-butene
Concen: 24.039 ug/l
RT: 12.469 min Scan# 1850
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01




Tgt Ion: 75 Resp: 16320
Ion Ratio Lower Upper
75 100
53 81.5 61.4 92.2
89 41.8 35.4 53.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024




#82
4-Chlorotoluene
Concen: 22.885 ug/l
RT: 12.945 min Scan# 1931
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17



Tgt Ion: 91 Resp: 164547
Ion Ratio Lower Upper
91 100
126 35.3 18.9 56.9





#83

tert-Butylbenzene

Concen: 21.854 ug/l

RT: 13.163 min Scan# 1968

Delta R.T. 0.000 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :

VD0808SBS01





Tgt Ion:119 Resp: 157138

Ion	Ratio	Lower	Upper
119	100		
91	61.3	28.2	84.7
134	27.8	13.4	40.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#84

1,2,4-Trimethylbenzene

Concen: 22.197 ug/l

RT: 13.210 min Scan# 1976




Delta R.T. 0.006 min


Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Tgt Ion:105 Resp: 187142

Ion	Ratio	Lower	Upper
105	100		
120	46.5	23.6	71.0





#85

sec-Butylbenzene

Concen: 22.337 ug/l

RT: 13.339 min Scan# 1998

Delta R.T. -0.000 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17

Instrument :

MSVOA_D

ClientSampleId :

VD0808SBS01

Tgt Ion:105 Resp: 245240

Ion Ratio Lower Upper

105 100





134 19.9 10.3 30.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024

Supervised By :Semsettin Yesilyurt 08/09/2024



#86

p-Isopropyltoluene

Concen: 21.626 ug/l

RT: 13.457 min Scan# 2018

Delta R.T. 0.000 min

Lab File: VD079667.D

Acq: 08 Aug 2024 15:17




Tgt Ion:119 Resp: 199852


Ion Ratio Lower Upper

119 100

134 26.5 13.7 41.1


91 22.6 10.7 32.1





#87
1,3-Dichlorobenzene
Concen: 21.039 ug/l
RT: 13.457 min Scan# 2031
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01



Tgt Ion:146 Resp: 108290
Ion Ratio Lower Upper
146 100
111 41.1 18.5 55.5
148 64.0 30.9 92.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#88
1,4-Dichlorobenzene
Concen: 21.641 ug/l
RT: 13.534 min Scan# 2031
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Tgt Ion:146 Resp: 109512
Ion Ratio Lower Upper
146 100
111 39.0 17.5 52.6
148 63.4 31.4 94.2









#89
n-Butylbenzene
Concen: 21.487 ug/l
RT: 13.787 min Scan# 2074
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01

Tgt Ion: 91 Resp: 183759
Ion Ratio Lower Upper
91 100
92 52.7 25.7 77.0
134 25.4 13.6 40.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#90
Hexachloroethane
Concen: 21.803 ug/l
RT: 14.051 min Scan# 2119
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Tgt Ion:117 Resp: 41871
Ion Ratio Lower Upper
117 100
201 66.6 36.3 108.9









#91
1,2-Dichlorobenzene
Concen: 21.797 ug/l
RT: 13.828 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17

Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01

Tgt Ion:146 Resp: 9588
Ion Ratio Lower Upper
146 100
111 41.9 19.1 57.1
148 65.0 31.1 93.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 25.376 ug/l
RT: 14.445 min Scan# 2186
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17


Tgt Ion: 75 Resp: 6500
Ion Ratio Lower Upper
75 100
155 81.2 47.7 143.1
157 102.1 63.4 190.2





#93
1,2,4-Trichlorobenzene
Concen: 19.378 ug/l
RT: 15.098 min Scan# 2166
Delta R.T. 0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01




Tgt Ion:180 Resp: 53708
Ion Ratio Lower Upper
180 100
182 93.6 46.9 140.7
145 30.0 14.0 41.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024




#94
Hexachlorobutadiene
Concen: 18.243 ug/l
RT: 15.198 min Scan# 2314
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Tgt Ion:225 Resp: 27289
Ion Ratio Lower Upper
225 100
223 63.7 33.3 99.9
227 66.4 33.6 100.6





#95
Naphthalene
Concen: 22.878 ug/l
RT: 15.328 min Scan# 21
Delta R.T. -0.000 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17




Instrument :
MSVOA_D
ClientSampleId :
VD0808SBS01




Tgt Ion:128 Resp: 10974
Ion Ratio Lower Upper
128 100
127 13.2 10.2 15.4
129 11.7 8.5 12.7

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024



#96
1,2,3-Trichlorobenzene
Concen: 20.118 ug/l
RT: 15.516 min Scan# 2368
Delta R.T. -0.006 min
Lab File: VD079667.D
Acq: 08 Aug 2024 15:17



Tgt Ion:180 Resp: 49048
Ion Ratio Lower Upper
180 100
182 90.9 47.8 143.3
145 31.6 16.1 48.2

