

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080924\
 Data File : VD079695.D
 Acq On : 09 Aug 2024 16:08
 Operator : RP/MD
 Sample : P3484-01RE
 Misc : 5.04G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BRE

Quant Time: Aug 10 02:11:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	4802	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.769	114	12417	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	12454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	5185	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	5900	120.942	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	241.880%#
35) Dibromofluoromethane	7.799	113	4621	54.983	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.960%
50) Toluene-d8	10.269	98	11972	36.682	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	73.360%
62) 4-Bromofluorobenzene	12.563	95	3975	42.088	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	84.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080924\
Data File : VD079695.D
Acq On : 09 Aug 2024 16:08
Operator : RP/MD
Sample : P3484-01RE
Misc : 5.04G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BRE

Quant Time: Aug 10 02:11:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

