

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080924\
 Data File : VD079698.D
 Acq On : 09 Aug 2024 17:31
 Operator : RP/MD
 Sample : P3489-02RE
 Misc : 5.08G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOCRE

Quant Time: Aug 10 02:12:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	34469	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	75831	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	55055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	12858	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	36219	103.432	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	206.860%#
35) Dibromofluoromethane	7.804	113	29447	57.373	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.740%
50) Toluene-d8	10.269	98	63433	31.825	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	63.660%
62) 4-Bromofluorobenzene	12.569	95	12488	21.651	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	43.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080924\
Data File : VD079698.D
Acq On : 09 Aug 2024 17:31
Operator : RP/MD
Sample : P3489-02RE
Misc : 5.08G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOCRE

Quant Time: Aug 10 02:12:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

