

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081224\
 Data File : VD079709.D
 Acq On : 12 Aug 2024 14:08
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Aug 13 01:28:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 13 01:06:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	97340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	474871	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	551937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	242959	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	122928	109.769	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	219.540%#
35) Dibromofluoromethane	7.804	113	192210	65.657	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	131.320%
50) Toluene-d8	10.269	98	679973	55.574	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	111.140%
62) 4-Bromofluorobenzene	12.569	95	226060	65.986	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	131.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081224\
Data File : VD079709.D
Acq On : 12 Aug 2024 14:08
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Aug 13 01:28:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081224S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 13 01:06:11 2024
Response via : Initial Calibration

