



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081424\
Data File : VD079727.D
Acq On : 14 Aug 2024 08:51
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081224S.M
Title : SW846 8260
Last Update : Tue Aug 13 01:06:11 2024



AutoFind: Scans 1866, 1867, 1868; Background Corrected with Scan 1858

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.3	13276	PASS
75	95	30	60	51.5	37408	PASS
95	95	100	100	100.0	72592	PASS
96	95	5	9	6.5	4730	PASS
173	174	0.00	2	1.9	1157	PASS
174	95	50	100	85.3	61952	PASS
175	174	5	9	8.5	5244	PASS
176	174	95	101	95.6	59237	PASS
177	176	5	9	7.6	4475	PASS