

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081820\
 Data File : VD066290.D
 Acq On : 18 Aug 2020 20:47
 Operator : VA/SY
 Sample : L3700-08
 Misc : 8.12G/5.00ml/MSVOA D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S408-K-SO-1-1.5-081420

Quant Time: Aug 19 08:04:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:49:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.97	168	187169	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	280915	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	225769	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	71301	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	100273	60.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.10%	
35) Dibromofluoromethane	7.91	113	96981	54.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.00%	
50) Toluene-d8	10.34	98	299119	45.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.10%	
62) 4-Bromofluorobenzene	12.64	95	76919	32.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.98%	
Target Compounds						
8) Diethyl Ether	3.71	74	2489	2.979	ug/l	89
16) Acetone	4.16	43	13801	43.559	ug/l	96
20) Methylene Chloride	4.96	84	24211	9.530	ug/l	95
25) 2-Butanone	7.21	43	3794	8.141	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081820\
Data File : VD066290.D
Acq On : 18 Aug 2020 20:47
Operator : VA/SY
Sample : L3700-08
Misc : 8.12G/5.00ml/MSVOA D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S408-K-SO-1-1.5-081420

Quant Time: Aug 19 08:04:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:49:25 2020
Response via : Initial Calibration

