

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081820\
 Data File : VD066307.D
 Acq On : 19 Aug 2020 06:44
 Operator : VA/SY
 Sample : L3712-02
 Misc : 8.34G/5.00ml/MSVOA D/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-27137

Quant Time: Aug 19 09:06:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:49:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.97	168	69911	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	123550	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	90646	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	31731	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	61144	98.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	196.06%	
35) Dibromofluoromethane	7.91	113	49449	63.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.36%	
50) Toluene-d8	10.34	98	93238	32.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.56%	
62) 4-Bromofluorobenzene	12.63	95	34255	32.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	65.80%	
Target Compounds						
8) Diethyl Ether	3.71	74	1207	3.868	ug/l	89
11) Tert butyl alcohol	5.21	59	9322	191.226	ug/l #	93
16) Acetone	4.16	43	23218	196.189	ug/l	95
20) Methylene Chloride	4.97	84	20920	28.803	ug/l	92
69) o-Xylene	12.18	106	1146	1.020	ug/l	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081820\
Data File : VD066307.D
Acq On : 19 Aug 2020 06:44
Operator : VA/SY
Sample : L3712-02
Misc : 8.34G/5.00ml/MSVOA D/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-27137

Quant Time: Aug 19 09:06:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:49:25 2020
Response via : Initial Calibration

